

## Report of Analysis

Client:	Kleinfelder	Date Collected:	12/23/24
Project:	Comegys School	Date Received:	12/23/24
Client Sample ID:	PIBLK-PL093481.D	SDG No.:	P5382
Lab Sample ID:	I.BLK-PL093481.D	Matrix:	WATER
Analytical Method:	SW8081	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group1
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL093481.D	1		12/23/24	PL122324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
309-00-2	Aldrin	0.0044	U	0.0044	0.050	ug/L
60-57-1	Dieldrin	0.0047	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.0045	U	0.0045	0.050	ug/L
72-54-8	4,4-DDD	0.0092	U	0.0092	0.050	ug/L
50-29-3	4,4-DDT	0.0044	U	0.0044	0.050	ug/L
<b>SURROGATES</b>						
2051-24-3	Decachlorobiphenyl	22.0		43 - 140	110%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.5		77 - 126	102%	SPK: 20

### Comments:

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 P = Indicates >25% difference for detected concentrations between the two GC columns  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.  
 () = Laboratory InHouse Limit