



Report of Analysis

Client: Date Collected: Kleinfelder 12/26/24 Project: Date Received: 12/26/24 Comegys School Client Sample ID: PIBLK-PL093525.D SDG No.: P5382 Lab Sample ID: I.BLK-PL093525.D Matrix: WATER

Analytical Method: SW8081 % Solid: 0 Decanted:

Sample Wt/Vol: 1000 Units: mL Final Vol: 10000 uL

Soil Aliquot Vol: uL Test: PESTICIDE Group1

Extraction Type: Injection Volume:

GPC Factor: 1.0 PH:

Prep Method: 3510C

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID
PL093525.D 1 1 12/26/24 pl122624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
309-00-2	Aldrin	0.0044	U	0.0044	0.050	ug/L
60-57-1	Dieldrin	0.0047	U	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.0045	U	0.0045	0.050	ug/L
72-54-8	4,4-DDD	0.0092	U	0.0092	0.050	ug/L
50-29-3	4,4-DDT	0.0044	U	0.0044	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	24.0		43 - 140	120%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.7		77 - 126	108%	SPK: 20

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit