



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5382

Client: Kleinfelder

Analytical Method: 8081B Datafile : PL093518.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165844BS	Aldrin	16.65	17.4	ug/kg	105				82	124	
	Dieldrin	16.65	18.7	ug/kg	112				85	121	
	4,4'-DDE	16.65	18.6	ug/kg	112				81	123	
	4,4'-DDD	16.65	19.2	ug/kg	115				80	131	
	4,4'-DDT	16.65	20.1	ug/kg	121				70	129	