

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG No.: P5382
 Instrument ID: MSVOA_Y Calibration Date(s): 12/26/2024 12/26/2024
 Heated Purge: (Y/N) Y Calibration Time(s): 09:31 11:51
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY020706.D		RRF010 = VY020707.D		RRF020 = VY020708.D		
		RRF050 = VY020709.D		RRF150 = VY020710.D		RRF100 = VY020711.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF150	RRF100	RRF	% RSD
cis-1,2-Dichloroethene	0.419	0.460	0.426	0.462	0.432	0.447	0.441	4.1
1,1,1-Trichloroethane	0.814	0.897	0.826	0.859	0.816	0.831	0.840	3.8
Benzene	0.969	1.104	1.040	1.089	1.041	1.083	1.054	4.7
Trichloroethene	0.288	0.336	0.315	0.324	0.309	0.317	0.315	5.1
Toluene	0.654	0.751	0.722	0.745	0.723	0.741	0.723	4.9
Ethyl Benzene	1.540	1.810	1.715	1.778	1.682	1.743	1.711	5.6
m/p-Xylenes	0.590	0.691	0.649	0.675	0.643	0.667	0.653	5.4
o-Xylene	0.552	0.666	0.639	0.650	0.603	0.632	0.624	6.6
Isopropylbenzene	3.123	3.801	3.508	3.476	3.402	3.474	3.464	6.3
1,2-Dichloroethane-d4	0.335	0.402	0.313	0.516	0.433	0.457	0.410	18.7
Dibromofluoromethane	0.235	0.274	0.232	0.319	0.284	0.307	0.275	13.1
Toluene-d8	0.733	0.883	0.733	1.227	1.093	1.158	0.971	22.4
4-Bromofluorobenzene	0.303	0.378	0.319	0.426	0.373	0.397	0.366	12.8

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.