

Surrogate Summary

SDG No.: **P5382**

Client: **Kleinfelder**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5382-01	COMP-1	1,2-Dichloroethane-d4	50	65.6	131	50	163
		Dibromofluoromethane	50	58.3	117	54	147
		Toluene-d8	50	53.0	106	58	134
		4-Bromofluorobenzene	50	50.7	101	29	146
P5382-02	COMP-2	1,2-Dichloroethane-d4	50	62.4	125	50	163
		Dibromofluoromethane	50	56.8	114	54	147
		Toluene-d8	50	53.0	106	58	134
		4-Bromofluorobenzene	50	44.5	89	29	146
P5382-03	COMP-3	1,2-Dichloroethane-d4	50	61.4	123	50	163
		Dibromofluoromethane	50	57.3	115	54	147
		Toluene-d8	50	52.6	105	58	134
		4-Bromofluorobenzene	50	47.0	94	29	146
VY1226SBL01	VY1226SBL01	1,2-Dichloroethane-d4	50	64.7	129	50	163
		Dibromofluoromethane	50	57.6	115	54	147
		Toluene-d8	50	52.6	105	58	134
		4-Bromofluorobenzene	50	51.0	102	29	146
VY1226SBS01	VY1226SBS01	1,2-Dichloroethane-d4	50	55.0	110	50	163
		Dibromofluoromethane	50	52.1	104	54	147
		Toluene-d8	50	50.6	101	58	134
		4-Bromofluorobenzene	50	51.8	104	29	146
VY1227SBL01	VY1227SBL01	1,2-Dichloroethane-d4	50	63.5	127	50	163
		Dibromofluoromethane	50	57.1	114	54	147
		Toluene-d8	50	53.3	107	58	134
		4-Bromofluorobenzene	50	48.3	97	29	146
VY1227SBS01	VY1227SBS01	1,2-Dichloroethane-d4	50	54.8	110	50	163
		Dibromofluoromethane	50	54.3	109	54	147
		Toluene-d8	50	52.6	105	58	134
		4-Bromofluorobenzene	50	53.1	106	29	146
VY1227SBSD01	VY1227SBSD01	1,2-Dichloroethane-d4	50	55.9	112	50	163
		Dibromofluoromethane	50	53.6	107	54	147
		Toluene-d8	50	52.8	105	58	134
		4-Bromofluorobenzene	50	53.2	106	29	146