

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382
Lab File ID: VY020709.D Date Analyzed: 12/26/2024
Instrument ID: MSVOA_Y Time Analyzed: 10:39
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	95919	7.72	134605	8.62	116249	11.42
UPPER LIMIT	191838	8.22	269210	9.122	232498	11.92
LOWER LIMIT	47959.5	7.22	67302.5	8.122	58124.5	10.92
EPA SAMPLE NO.						
COMP-1	97104	7.71	158398	8.62	137952	11.42
VY1226SBL01	99549	7.72	157271	8.62	133794	11.42
VY1226SBS01	120044	7.72	165739	8.62	139363	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.