

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: P5382 SAS No.: P5382 SDG NO.: P5382
 Lab File ID: VY020709.D Date Analyzed: 12/26/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	60351	13.353				
UPPER LIMIT	120702	13.853				
LOWER LIMIT	30175.5	12.853				
EPA SAMPLE NO.						
COMP-1	58207	13.35				
VY1226SBL01	56315	13.35				
VY1226SBS01	74040	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.