

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID:	P5385	OrderDate:	12/26/2024 12:37:00 PM
Client:	Aramark Uniforms	Project:	Monthly 2024
Contact:	Jose Liceaga	Location:	L63

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5385-01	GRAB	WATER			12/26/24 12:30			12/26/24
			TPH	1664A			12/27/24 11:25	
P5385-02	COMP	WATER			12/26/24 12:34			12/26/24
			BOD5	SM5210 B			12/26/24 16:40	
			TSS	SM2540 D			12/27/24 08:15	



SAMPLE DATA

Report of Analysis

Client:	Aramark Uniforms	Date Collected:	12/26/24 12:30
Project:	Monthly 2024	Date Received:	12/26/24
Client Sample ID:	GRAB	SDG No.:	P5385
Lab Sample ID:	P5385-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TPH	41.8		1	0.40	5.00	mg/L		12/27/24 11:25	1664A

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Aramark Uniforms	Date Collected:	12/26/24 12:34
Project:	Monthly 2024	Date Received:	12/26/24
Client Sample ID:	COMP	SDG No.:	P5385
Lab Sample ID:	P5385-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
BOD5	1180		1	0.17	2.00	mg/L		12/26/24 16:40	SM 5210 B-16
TSS	945		1	1.00	4.00	mg/L		12/27/24 08:15	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Preparation Blank Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: BOD5	LB134094BL mg/L	< 0.2000	0.2000	U	0.17	2.0	12/26/2024
Sample ID: TSS	LB134098BL mg/L	< 2.0000	2.0000	U	1	4	12/27/2024
Sample ID: TPH	LB134104BL mg/L	< 2.5000	2.5000	U	0.4	5.0	12/27/2024

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: P5385
Project: Monthly 2024	Sample ID: LB134104BS
Client ID: LB134104BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.7		17.0		1	1.78		12/27/2024

Duplicate Sample Summary

Client:	Aramark Uniforms	SDG No.:	P5385
Project:	Monthly 2024	Sample ID:	P5385-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	1180		1170		1	0.26		12/26/2024
TSS	mg/L	+/-5	945		949		1	0.42		12/27/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134094

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134094BS							
BOD5	mg/L	198	171		86	1	84.6-115.4	12/26/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134098

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134098BS							
TSS	mg/L	550	510		93	1	90-110	12/27/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134104BS							
TPH	mg/L	20.0	16.7		84	1	78-114	12/27/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134104BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	12/27/2024



RAW DATA

BOD5 LOG

ANALYST: rubin
 On:12/31/2024 1:56:12 PM

SUPERVISOR: Iwona

QC BATCH ID: LB134094

Analysis Date: 12/26/2024

BOD Water: WP111235

MANGANOUS SULFATE SOLUTION: W3103

Starch: W3149

Alkaline Iodide Azide: W3109

Sulfuric acid, 1N: WP110386

Sodium Thiosulfate, 0.025N: W3105

POLYSEED: WP111237

NaOH, 1N: WP108662

GGA: WP111236

IncubatorID: INCUBATOR #3

Chlorine Strips: W3155

GuageID: 0511062

pH Strips: W3140

Zero DO: WP111005

Lab SampleID	Client ID	Bottle No.	VOL. ML	Initial Reading (ML)	Final Reading (ML)	Difference	Average
WINKLER 1	WINKLER 1	1	300	0.0	10.01	10.01	10.01
WINKLER 2	WINKLER 2	2	300	10.00	20.01	10.01	10.01

Meter Calibration1: 9.89 Zero DO Reading1: 0.10 mg/L (<=0.2 Criteria)

Barometric Pressure1: 771 mmHg DO Meter BOD fluid reading for winkler comparison: 10.17

After Incubation

Meter Calibration2: 9.28 Zero DO Reading2: 0.10 mg/L (<=0.2 Criteria)

Barometric Pressure2: 755 mmHg



QC BATCH ID: LB134094

INCUBATOR TEMP IN(C): 19.8

INCUBATOR TEMP OUT(C): 19.9

TIME IN: 16:40

TIME OUT: 13:00

DATE IN: 12/26/2024

DATE OUT: 12/31/2024

Lab SampleID	Bottle No.	Check CL	Initial PH	Final PH	Temp °C	Sam Vol. (mL)	D.O.1 Initial	D.O.2 Final	Depletion	BOD Result (mg/L)	Avg Result (mg/L)	Comment
LB134094BL	1	No	6.66	N/A	20.70	300	10.16	10.14	0.02	0.02	0.02	
POLYSEED	1					10	10.13	6.91	3.22	0.64	0.64	
POLYSEED	2					15	10.10	5.11	4.99	0.67		
POLYSEED	3					20	10.02	3.81	6.21	0.62		
GGA	1					6	10.11	6.11	4	168	171.17	
GGA	2					6	10.07	6.02	4.05	170.5		
GGA	3					6	10.03	5.89	4.14	175		
P5385-02	1	No	6.04	6.90	20.00	0.5	10.04	7.69	2.35	1026	1175.5	pH Adjusted
P5385-02	2					1	10.02	4.82	5.2	1368		
P5385-02	3					2	9.85	1.66	8.19	1132.5		
P5385-02	4					3	9.72	0.06	-	0		
P5385-02DUP	1	No	6.04	6.90	20.00	0.5	10.05	7.35	2.7	1236	1172.5	pH Adjusted
P5385-02DUP	2					1	10.00	5.61	4.39	1125		
P5385-02DUP	3					2	9.85	1.50	8.35	1156.5		
P5385-02DUP	4					3	9.72	0.06	-	0		
P5386-05	1	No	7.40	N/A	20.10	5	9.97	8.50	-	0	55.89	
P5386-05	2					20	9.95	4.69	5.26	69.3		
P5386-05	3					50	9.91	2.19	7.72	42.48		
P5386-05	4					150	9.70	0.29	-	0		

NOTE: 2ml POLYSEED added to GGA and all the Samples, but not in Blank.

NOTE (For, CBOD5): 0.16 g Nitrification Inhibitor added to GGA and all the Samples, but not in Blank.

WORKLIST(Hardcopy Internal Chain)

LB134094

WorkList Name : BOD5-12-26. WorkList ID : 186627 Department : Wet-Chemistry Date : 12-26-2024 13:13:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5385-02 A	COMP	Water	BOD5	Cool 4 deg C	ARAM01	L63	12/26/2024	SM5210 B
P5386-05 R	MOO-24-00397	Water	BOD5	Cool 4 deg C	PSEG03	N31	12/26/2024	SM5210 B

Date/Time 12/26/2024 15:00
Raw Sample Received by: RM CWL
Raw Sample Relinquished by: RM CWL

Date/Time 12/26/2024 16:30
Raw Sample Received by: RM CWL
Raw Sample Relinquished by: RM CWL

TOTAL SUSPENDED SOLIDS - SM2540D

TEMP1 IN: 103 °C 12/26/2024 14:00 **TEMP1 OUT:** 104 °C 12/26/2024 15:00
TEMP2 IN: 103 °C 12/26/2024 15:30 **TEMP2 OUT:** 104 °C 12/26/2024 16:30
TEMP3 IN: 103 °C 12/27/2024 08:15 **TEMP3 OUT:** 104 °C 12/27/2024 09:45
TEMP4 IN: 104 °C 12/27/2024 10:15 **TEMP4 OUT:** 104 °C 12/27/2024 11:45

SUPERVISOR: Iwona

ANALYST: Niha

Date: 12/26/2024

Run Number: LB134098

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB134098BL	LB134098BL	1.4695	1.4695	100	1.4695	1.4695	1.4695	0.0000	0
2	LB134098BS	LB134098BS	1.4520	1.4520	100	1.5030	1.5030	1.5030	0.0510	510
3	P5385-02	COMP	1.4840	1.4840	100	1.5785	1.5785	1.5785	0.0945	945
4	P5385-02DUP	COMPDUP	1.4637	1.4637	100	1.5586	1.5586	1.5586	0.0949	949
5	P5386-05	MOO-24-00397	1.4720	1.4720	100	1.7483	1.7483	1.7483	0.2763	2763

A = Sample Volume (ml)
B = Final Empty Dish Weight (g)
C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB134098

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS-12262024 WorkList ID : 186630 Department : Wet-Chemistry Date : 12-26-2024 15:16:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5385-02	COMP	Water	TSS	Cool 4 deg C	ARAM01	L63	12/26/2024	SM2540 D
P5386-05	MOO-24-00397	Water	TSS	Cool 4 deg C	PSEG03	N31	12/26/2024	SM2540 D

Date/Time 12.27.2024, 07:45
Raw Sample Received by: NFWC
Raw Sample Relinquished by: NFWC

Date/Time 12.27.2024, 09:10
Raw Sample Received by: NFWC
Raw Sample Relinquished by: NFWC

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB134104
Analysis Date: 12/27/2024
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 12/27/2024
Extraction IN Time: 09:50
Extraction OUT Time: 10:35
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB134104BL	LB134104BL	WATER	1.3	1000	100	2.5639	2.5639	3.01	2.5640	2.5640	0.0001	0.1
2	LB134104BS	LB134104BS	WATER	1.3	1000	100	2.8963	2.8963	3.02	2.9130	2.9130	0.0167	16.7
3	LB134104BSD	LB134104BSD	WATER	1.3	1000	100	2.9971	2.9971	3.02	3.0141	3.0141	0.0170	17
4	P5385-01	GRAB	WATER	1.6	1000	100	3.0087	3.0087	3.05	3.0505	3.0505	0.0418	41.8
5	P5386-05	MOO-24-00397	WATER	1.6	1000	100	3.0314	3.0314	3.04	3.0700	3.0700	0.0386	38.6



Alliance
TECHNICAL GROUP

Test: TPH

Analysis Date: 12/27/2024

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3153
pH Paper 0-14	M6069
Sodium Sulfate	EP2573
1:1 HCL	WP110826
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP100827
LCSWD	5.00 ML	WP100828
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 12:16

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 11:25

Bal Check Time: 10:10 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 13:00

Out Time1: 12:15

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 14:11

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 13:30

Bal Check Time: 14:40 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 14:37

```
Out Time2:      14:10
```

WORKLIST(Hardcopy Internal Chain)

LB 134104

WorkList Name : tph p5377

WorkList ID : 186661

Department : Wet-Chemistry

Date : 12-27-2024 09:25:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5385-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	L63	12/26/2024	1664A
P5386-05	MOO-24-00397	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	N31	12/26/2024	1664A

Date/Time 12/27/24 09:13:30

Raw Sample Received by: JB Lueck

Raw Sample Relinquished by: CB 502

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Instrument ID: DO METER

Daily Analysis Runlog For Sequence/QC Batch ID # LB134094

Review By	rubina	Review On	12/31/2024 1:55:52 PM
Supervise By	Iwona	Supervise On	12/31/2024 1:56:12 PM
SubDirectory	LB134094	Test	BOD5
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP111235,W3149,WP110386,W3103,W3109,W3105,WP111237,WP111236,WP108662		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB134094BL	LB134094BL	MB	12/26/24 16:40		rubina	OK
2	LB134094BS	LB134094BS	LCS	12/26/24 16:40		rubina	OK
3	P5385-02	COMP	SAM	12/26/24 16:40		rubina	OK
4	P5385-02DUP	COMPDUP	DUP	12/26/24 16:40		rubina	OK
5	P5386-05	MOO-24-00397	SAM	12/26/24 16:40		rubina	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB134098

Review By	Niha	Review On	12/27/2024 12:05:36 PM
Supervise By	Iwona	Supervise On	12/27/2024 1:06:32 PM
SubDirectory	LB134098	Test	TSS
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB134098BL	LB134098BL	MB	12/27/24 08:15		Niha	OK
2	LB134098BS	LB134098BS	LCS	12/27/24 08:15		Niha	OK
3	P5385-02	COMP	SAM	12/27/24 08:15		Niha	OK
4	P5385-02DUP	COMPDUP	DUP	12/27/24 08:15		Niha	OK
5	P5386-05	MOO-24-00397	SAM	12/27/24 08:15		Niha	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB134104

Review By	jignesh	Review On	12/27/2024 10:07:32 AM
Supervise By	Iwona	Supervise On	12/27/2024 10:57:04 AM
SubDirectory	LB134104	Test	TPH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3153,M6069,EP2573,WP110826,W3079,NA,WP100827,WP100828,NA		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB134104BL	LB134104BL	MB	12/27/24 11:25		jignesh	OK
2	LB134104BS	LB134104BS	LCS	12/27/24 11:25		jignesh	OK
3	LB134104BSD	LB134104BSD	LCSD	12/27/24 11:25		jignesh	OK
4	P5385-01	GRAB	SAM	12/27/24 11:25		jignesh	OK
5	P5386-05	MOO-24-00397	SAM	12/27/24 11:25		jignesh	OK



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. **P53 85**
QUOTE NO.
COC Number **2041575**

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: **Aramark Uniforms**
ADDRESS: **740 Frelinghuysen Ave**
CITY: **Newark** STATE: **NJ** ZIP: **07114**
ATTENTION: **Jarrod Mills**
PHONE: **973-824-1101** FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **Monthly**
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1 2 3 4 5 6 7 8 9
TPH DOBS ISS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	Grab Comp	W		✓	12-26-24	1230	1	✓									← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
2.		W	✓		12-26-24	1234	2	✓	✓								
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1.	DATE/TIME: 1237 12-26-24	RECEIVED BY: 1.	DATE/TIME: 1237 12-26-24	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 4.4° °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME: 1406	RECEIVED BY: 2.	DATE/TIME: 1406	Comments:
RELINQUISHED BY SAMPLER: 3.	DATE/TIME: 1606 12-26-24	RECEIVED BY: 3.	DATE/TIME: 1606	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete
☐ YES ☐ NO



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488