

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134094

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134094BS							
BOD5	mg/L	198	171		86	1	84.6-115.4	12/26/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134098

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134098BS							
TSS	mg/L	550	510		93	1	90-110	12/27/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134104BS							
TPH	mg/L	20.0	16.7		84	1	78-114	12/27/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5385

Project: Monthly 2024

Run No.: LB134104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134104BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	12/27/2024