

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51811 MA No. : SDG No.: MH2D26
 Initial Calibration Verification Source : MP83957
 Continuing Calibration Verification Source : MP83959
 Run Batch: LB134207 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV050				ID: CCV020				ID: CCV021		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	95		5.0	5.0	101		4.9	99	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51811 MA No. : SDG No.: MH2D26
Initial Calibration Verification Source : MP83957
Continuing Calibration Verification Source : MP83959
Run Batch: LB134207 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV022				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	4.9	98				