


LB134207

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>9/01/2025</i>
0.00	0.000	0.092	0.092	354	0.000	354	0				
0.05	0.050					442	0				
0.20	0.200	0.215	0.015	1065	0.0 %	1065					
2.50	2.500	2.586	0.086	14753	0.0 %	14753					
5.00	5.000	4.862	-0.138	27892	0.0 %	27892					
7.50	7.500	7.179	-0.321	41266	0.0 %	41266					
10.0	10.000	10.267	0.267	59098	0.0 %	59098					3

LB134207
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	354	-		0 SFAM01.1	PPB	1/9/2025 10:00	Std
	0.2 S01	1065	-		0.2 SFAM01.1	PPB	1/9/2025 10:02	Std
	2.5 S02	14753	-		2.5 SFAM01.1	PPB	1/9/2025 10:04	Std
	5 S03	27892	-		5 SFAM01.1	PPB	1/9/2025 10:07	Std
	7.5 S04	41266	-		7.5 SFAM01.1	PPB	1/9/2025 10:12	Std
	10 S05	59098	-		10 SFAM01.1	PPB	1/9/2025 10:19	Std
ICV050	ICV050	21832	3.8121 -		SFAM01.1	PPB	1/9/2025 10:28	SMPL
ICB050	ICB050	344	0.0899 -		SFAM01.1	PPB	1/9/2025 10:30	SMPL
CCV020	CCV020	28947	5.0446 -		SFAM01.1	PPB	1/9/2025 10:32	SMPL
CCB020	CCB020	257	0.0748 -		SFAM01.1	PPB	1/9/2025 10:34	SMPL
PB165987BL	PBS987	-3	0.0298 -		SFAM01.1	PPB	1/9/2025 10:37	SMPL
P5388-01	MH2D26	2575	0.4763 -		SFAM01.1	PPB	1/9/2025 10:39	SMPL
P5388-02	MH2D29	5376	0.9615 -		SFAM01.1	PPB	1/9/2025 10:41	SMPL
P5388-03	MH2D32	2317	0.4316 -		SFAM01.1	PPB	1/9/2025 10:43	SMPL
P5388-04	MH2D35	5255	0.9406 -		SFAM01.1	PPB	1/9/2025 10:46	SMPL
P5388-05	MH2D38	13348	2.3425 -		SFAM01.1	PPB	1/9/2025 10:48	SMPL
P5388-06	MH2D41	1173	0.2335 -		SFAM01.1	PPB	1/9/2025 10:50	SMPL
P5388-07	MH2D44	52590	9.1401 -		SFAM01.1	PPB	1/9/2025 10:53	SMPL
P5388-08	MH2D47	28225	4.9195 -		SFAM01.1	PPB	1/9/2025 10:55	SMPL
P5388-09	MH2D50	47980	8.3415 -		SFAM01.1	PPB	1/9/2025 10:57	SMPL
P5388-10	MH2D53	1479	0.2865 -		SFAM01.1	PPB	1/9/2025 10:59	SMPL
P5388-11	MH2D56	81197	14.0955 -		SFAM01.1	PPB	1/9/2025 11:02	SMPL
P5388-12	MH2D59	1105	0.2217 -		SFAM01.1	PPB	1/9/2025 11:04	SMPL
P5388-13	MH2D62	539	0.1236 -		SFAM01.1	PPB	1/9/2025 11:06	SMPL
P5388-14	MH2D65	825	0.1732 -		SFAM01.1	PPB	1/9/2025 11:09	SMPL
P5388-15	MH2D68	1544	0.2977 -		SFAM01.1	PPB	1/9/2025 11:11	SMPL
P5388-16	MH2D71	1028	0.2084 -		SFAM01.1	PPB	1/9/2025 11:13	SMPL
P5388-17	MH2D74	783	0.1659 -		SFAM01.1	PPB	1/9/2025 11:15	SMPL
P5388-18	MH2D74D	763	0.1624 -		SFAM01.1	PPB	1/9/2025 11:18	SMPL
P5388-19	MH2D74S	12152	2.1353 -		SFAM01.1	PPB	1/9/2025 11:20	SMPL
CCV021	CCV021	28370	4.9446 -		SFAM01.1	PPB	1/9/2025 11:22	SMPL
CCB021	CCB021	247	0.0731 -		SFAM01.1	PPB	1/9/2025 11:24	SMPL
P5388-11DLX5	MH2D56	16435	2.8772 -		SFAM01.1	PPB	1/9/2025 11:27	SMPL
CCV022	CCV022	28112	4.8999 -		SFAM01.1	PPB	1/9/2025 11:29	SMPL
CCB022	CCB022	259	0.0751 -		SFAM01.1	PPB	1/9/2025 11:31	SMPL