

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134172

Review By	Sarabjit Jaswal	Review On	1/10/2025 12:10:23 PM
Supervise By	Mohan Bera	Supervise On	1/16/2025 1:30:08 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	01/06/25 12:39		Jaswal	OK
2	S00	S0	CAL1	01/06/25 13:15		Jaswal	OK
3	S02	S02	CAL3	01/06/25 13:19		Jaswal	OK
4	S03	S03	CAL4	01/06/25 13:25		Jaswal	OK
5	S04	S04	CAL5	01/06/25 13:28		Jaswal	OK
6	S05	S05	CAL6	01/06/25 13:31		Jaswal	OK
7	S06	S06	CAL7	01/06/25 13:34		Jaswal	OK
8	S07	S07	CAL8	01/06/25 13:36		Jaswal	OK
9	S08	S08	CAL9	01/06/25 13:39		Jaswal	OK
10	ICV001	ICV001	ICV	01/06/25 14:05	Inst. paused for calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	01/06/25 14:24	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	01/06/25 14:27		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	01/06/25 14:31		Jaswal	OK
14	CCV001	CCV001	CCV	01/06/25 14:34		Jaswal	OK
15	CCB001	CCB001	CCB	01/06/25 14:41	Inst. paused for review	Jaswal	OK
16	PB165957BL	PBS957	MB	01/06/25 14:58	Not Use	Jaswal	Not Ok
17	PB165957BS	LCS957	LCS	01/06/25 15:01	Not Use	Jaswal	Not Ok
18	P5365-01DL	TAPFTA-SB011-4.5-12	SAM	01/06/25 15:03	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134172

Review By	Sarabjit Jaswal	Review On	1/10/2025 12:10:23 PM
Supervise By	Mohan Bera	Supervise On	1/16/2025 1:30:08 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

19	P5365-01DUPDL	TAPFTA-SB01I-4.5-12	DUP	01/06/25 15:07	Not Use	Jaswal	Not Ok
20	P5365-01LDL	TAPFTA-SB01I-4.5-12	SD	01/06/25 15:10	Not Use	Jaswal	Not Ok
21	P5365-01DLMS	TAPFTA-SB01I-4.5-12	SAM	01/06/25 15:13	Not Use	Jaswal	Not Ok
22	P5365-01DLMSD	TAPFTA-SB01I-4.5-12	SAM	01/06/25 15:16	Not Use	Jaswal	Not Ok
23	P5365-01ADL	TAPFTA-SB01I-4.5-12	PS	01/06/25 15:19	Not Use	Jaswal	Not Ok
24	CCV002	CCV002	CCV	01/06/25 15:35		Jaswal	OK
25	CCB002	CCB002	CCB	01/06/25 15:38		Jaswal	OK
26	PB165924BL	PBS924	MB	01/06/25 15:41		Jaswal	OK
27	PB165924BS	LCS924	LCS	01/06/25 15:44		Jaswal	OK
28	P5388-01	MH2D26	SAM	01/06/25 15:48	Pb,Mn High	Jaswal	Dilution
29	P5388-02	MH2D29	SAM	01/06/25 15:51	Pb,Mn,Zn High , INT_STD 209Bi(1) Fail	Jaswal	Dilution
30	P5388-03	MH2D32	SAM	01/06/25 15:54	Pb High	Jaswal	Dilution
31	P5388-04	MH2D35	SAM	01/06/25 15:57	Pb,Mn High, INT_STD 209Bi(1) Fail	Jaswal	Dilution
32	P5388-05	MH2D38	SAM	01/06/25 16:00	As,Pb High, INT_STD 209Bi(1) Fail	Jaswal	Dilution
33	P5388-06	MH2D41	SAM	01/06/25 16:03	Pb High	Jaswal	Dilution
34	P5388-07	MH2D44	SAM	01/06/25 16:06	As,Pb High , INT_STD 209Bi(1) Fail	Jaswal	Dilution
35	P5388-08	MH2D47	SAM	01/06/25 16:09	As,Pb,Zn High, INT_STD 209Bi(1) Fail	Jaswal	Dilution
36	P5388-09	MH2D50	SAM	01/06/25 16:12	Pb High, INT_STD 209Bi(1) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134172

Review By	Sarabjit Jaswal	Review On	1/10/2025 12:10:23 PM
Supervise By	Mohan Bera	Supervise On	1/16/2025 1:30:08 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

37	P5388-10	MH2D53	SAM	01/06/25 16:15	Pb high	Jaswal	Dilution
38	P5388-11	MH2D56	SAM	01/06/25 16:18	As,Cu,Pb High, INT_STD 209Bi(1) Fail	Jaswal	Dilution
39	P5388-12	MH2D59	SAM	01/06/25 16:21		Jaswal	OK
40	P5388-13	MH2D62	SAM	01/06/25 16:24		Jaswal	OK
41	P5388-14	MH2D65	SAM	01/06/25 16:27	Mn High	Jaswal	Dilution
42	P5388-15	MH2D68	SAM	01/06/25 16:31	Pb,Mn,Zn High	Jaswal	Dilution
43	P5388-16	MH2D71	SAM	01/06/25 16:34		Jaswal	OK
44	P5388-17	MH2D74	SAM	01/06/25 16:37		Jaswal	OK
45	P5388-18	MH2D74D	DUP	01/06/25 16:40		Jaswal	OK
46	P5388-19	MH2D74S	MS	01/06/25 16:43		Jaswal	OK
47	P5388-17L	MH2D74L	SD	01/06/25 16:46		Jaswal	OK
48	CCV003	CCV003	CCV	01/06/25 16:49		Jaswal	OK
49	CCB003	CCB003	CCB	01/06/25 16:52		Jaswal	OK
50	P5388-01DL	MH2D26	SAM	01/06/25 17:48	2X For Pb,Mn Still Mn High	Jaswal	Dilution
51	P5388-02DL	MH2D29	SAM	01/06/25 17:51	2X For Pb,Mn,Zn, still Pb,Mn High, 2X for INT_STD 209Bi(1) Still Fail so report From Original	Jaswal	Dilution
52	P5388-04DL	MH2D35	SAM	01/06/25 17:54	2X for Pb,Mn,Still High, 2X for INT_STD 209Bi(1) Still Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134172

Review By	Sarabjit Jaswal	Review On	1/10/2025 12:10:23 PM
Supervise By	Mohan Bera	Supervise On	1/16/2025 1:30:08 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

53	P5388-05DL	MH2D38	SAM	01/06/25 17:58	2X For INT_STD 209Bi(1), Still Fail So Report From Original, 2X For As, Pb, Still Pb High	Jaswal	Dilution
54	P5388-06DL	MH2D41	SAM	01/06/25 18:01	2X For Pb	Jaswal	Confirms
55	P5388-07DL	MH2D44	SAM	01/06/25 18:04	2X For As, Pb, Still High, 2X For INT_STD 209Bi(1), Still Fail So Report From Original	Jaswal	Dilution
56	P5388-08DL	MH2D47	SAM	01/06/25 18:07	INT_STD 209Bi(1) Fail, 2X For As, Pb, Zn, Still Pb High	Jaswal	Dilution
57	P5388-09DL	MH2D50	SAM	01/06/25 18:10	2X For Pb, Still High, INT_STD 209Bi(1) Fail, So Report From Original	Jaswal	Dilution
58	P5388-10DL	MH2D53	SAM	01/06/25 18:13	2X For Pb	Jaswal	Confirms
59	P5388-11DL	MH2D56	SAM	01/06/25 18:16	2X For INT_STD 209Bi(1) Still Fail, 2X For As, Cu, Pb, Still As, Pb High	Jaswal	Dilution
60	P5388-03DL	MH2D32	SAM	01/06/25 18:19	4X For Pb	Jaswal	Confirms
61	P5388-04DL2	MH2D35	SAM	01/06/25 18:23	4X For Pb, Mn, Still Pb High, INT_STD 209Bi(1) Fail	Jaswal	Dilution
62	P5388-05DL2	MH2D38	SAM	01/06/25 18:26	4X For Pb	Jaswal	Confirms
63	CCV004	CCV004	CCV	01/06/25 18:29		Jaswal	OK
64	CCB004	CCB004	CCB	01/06/25 18:32		Jaswal	OK
65	P5388-01DL2	MH2D26	SAM	01/06/25 18:35	5X For Mn	Jaswal	Confirms
66	P5388-02DL2	MH2D29	SAM	01/06/25 18:38	5X For Pb, Mn, Still Pb High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134172

Review By	Sarabjit Jaswal	Review On	1/10/2025 12:10:23 PM
Supervise By	Mohan Bera	Supervise On	1/16/2025 1:30:08 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

67	P5388-07DL2	MH2D44	SAM	01/06/25 18:41	5X For As,Pb , Still Pb High , INT_STD 209Bi(1) Fail	Jaswal	Dilution
68	P5388-08DL2	MH2D47	SAM	01/06/25 18:44	Not Use	Jaswal	Not Ok
69	P5388-09DL2	MH2D50	SAM	01/06/25 18:47	Not Use	Jaswal	Not Ok
70	P5388-11DL2	MH2D56	SAM	01/06/25 18:51	INT_STD 209Bi(1) Fail,5X For As,Pb,Still Pb High	Jaswal	Dilution
71	P5388-14DL	MH2D65	SAM	01/06/25 18:54	5X For Mn	Jaswal	Confirms
72	P5388-15DL	MH2D68	SAM	01/06/25 18:57	5X For Pb,Mn,Zn	Jaswal	Confirms
73	P5388-04DL3	MH2D35	SAM	01/06/25 19:00	8X For Pb , 8X For INT_STD 209Bi(1) Still Fail	Jaswal	Dilution
74	P5388-02DL3	MH2D29	SAM	01/06/25 19:03	10X For Pb	Jaswal	Confirms
75	P5388-07DL3	MH2D44	SAM	01/06/25 19:09	10X For Pb , 10X For INT_STD 209Bi(1) Still Fail	Jaswal	Dilution
76	P5388-08DL3	MH2D47	SAM	01/06/25 19:12	10X For Pb , 10X For INT_STD 209Bi(1) Still Fail	Jaswal	Dilution
77	P5388-09DL3	MH2D50	SAM	01/06/25 19:15	10X For Pb	Jaswal	Confirms
78	P5388-11DL3	MH2D56	SAM	01/06/25 19:19	20X For Pb , 20X For INT_STD 209Bi(1) Still fail	Jaswal	Dilution
79	CCV005	CCV005	CCV	01/06/25 19:22		Jaswal	OK
80	CCB005	CCB005	CCB	01/06/25 19:25		Jaswal	OK