

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51811 MA No. : SDG No.: MH2D27
Initial Calibration Verification Source : MP83861
Continuing Calibration Verification Source : MP83863
Run Batch: LB134154 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV041				ID: CCV096				ID: CCV097		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.9	98		5.0	5.0	100		5.1	103	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51811 MA No. : SDG No.: MH2D27
 Initial Calibration Verification Source : MP83861
 Continuing Calibration Verification Source : MP83863
 Run Batch: LB134154 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV098				ID: CCV099		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.2	103		5.1	103	