


LB134154

SFAM01.1

INSTRUMENT ID:CV1

Linear



A= 0.0000e+000
B= 2.3315e-004 *slope*
C= -6.8597e-002 *= intercept*
Rho= 0.9998685
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.035	-0.035	145	0.000	145	0				
0.05	0.050					442	0				
0.20	0.200	0.203	0.003	1164	0.0 %	1164					
2.50	2.500	2.542	0.042	11195	0.0 %	11195					
5.00	5.000	4.953	-0.047	21536	0.0 %	21536					
7.50	7.500	7.607	0.107	32922	0.0 %	32922					
10.0	10.000	9.931	-0.069	42888	0.0 %	42888					

LBN134154
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	145	-		0 SFAM01.1	PPB	1/3/2025 10:39	Std
	0.2 S01	1164	-		0.2 SFAM01.1	PPB	1/3/2025 10:41	Std
	2.5 S02	11195	-		2.5 SFAM01.1	PPB	1/3/2025 10:43	Std
	5 S03	21536	-		5 SFAM01.1	PPB	1/3/2025 10:46	Std
	7.5 S04	32922	-		7.5 SFAM01.1	PPB	1/3/2025 10:51	Std
	10 S05	42888	-		10 SFAM01.1	PPB	1/3/2025 10:56	Std
ICV041	ICV041	17183	3.9376 -		SFAM01.1	PPB	1/3/2025 10:59	SMPL
ICB041	ICB041	56	-0.0555 -		SFAM01.1	PPB	1/3/2025 11:01	SMPL
CCV096	CCV096	21829	5.0208 -		SFAM01.1	PPB	1/3/2025 11:03	SMPL
CCB096	CCB096	66	-0.0532 -		SFAM01.1	PPB	1/3/2025 11:05	SMPL
PB165948BL	PBW948	158	-0.0318 -		SFAM01.1	PPB	1/3/2025 11:08	SMPL
P5392-01	MH2D27	1438	0.2667 -		SFAM01.1	PPB	1/3/2025 11:10	SMPL
P5392-02	MH2D30	3826	0.8234 -		SFAM01.1	PPB	1/3/2025 11:12	SMPL
P5392-03	MH2D33	634	0.0792 -		SFAM01.1	PPB	1/3/2025 11:14	SMPL
P5392-04	MH2D36	1380	0.2532 -		SFAM01.1	PPB	1/3/2025 11:17	SMPL
P5392-05	MH2D39	236	-0.0136 -		SFAM01.1	PPB	1/3/2025 11:19	SMPL
P5392-06	MH2D42	1374	0.2518 -		SFAM01.1	PPB	1/3/2025 11:21	SMPL
P5392-07	MH2D45	491	0.0459 -		SFAM01.1	PPB	1/3/2025 11:24	SMPL
P5392-08	MH2D48	659	0.085 -		SFAM01.1	PPB	1/3/2025 11:26	SMPL
P5392-09	MH2D51	911	0.1438 -		SFAM01.1	PPB	1/3/2025 11:28	SMPL
P5392-10	MH2D54	186	-0.0252 -		SFAM01.1	PPB	1/3/2025 11:30	SMPL
P5392-11	MH2D57	1237	0.2198 -		SFAM01.1	PPB	1/3/2025 11:33	SMPL
P5392-12	MH2D60	151	-0.0334 -		SFAM01.1	PPB	1/3/2025 11:35	SMPL
P5392-13	MH2D63	186	-0.0252 -		SFAM01.1	PPB	1/3/2025 11:37	SMPL
P5392-14	MH2D66	274	-0.0047 -		SFAM01.1	PPB	1/3/2025 11:42	SMPL
P5392-15	MH2D69	112	-0.0425 -		SFAM01.1	PPB	1/3/2025 11:44	SMPL
P5392-16	MH2D72	442	0.0345 -		SFAM01.1	PPB	1/3/2025 11:46	SMPL
P5392-17	MH2D75	184	-0.0257 -		SFAM01.1	PPB	1/3/2025 11:48	SMPL
P5392-18	MH2D75D	224	-0.0164 -		SFAM01.1	PPB	1/3/2025 11:51	SMPL
P5392-19	MH2D75S	5039	1.1062 -		SFAM01.1	PPB	1/3/2025 11:53	SMPL
CCV097	CCV097	22369	5.1467 -		SFAM01.1	PPB	1/3/2025 11:55	SMPL
CCB097	CCB097	170	-0.029 -		SFAM01.1	PPB	1/3/2025 11:58	SMPL
PB165949BL	PBW949	180	-0.0266 -		SFAM01.1	PPB	1/3/2025 12:00	SMPL
P5393-01	MH2D28	461	0.0389 -		SFAM01.1	PPB	1/3/2025 12:02	SMPL
P5393-02	MH2D31	602	0.0718 -		SFAM01.1	PPB	1/3/2025 12:04	SMPL
P5393-03	MH2D34	387	0.0216 -		SFAM01.1	PPB	1/3/2025 12:07	SMPL
P5393-04	MH2D37	490	0.0456 -		SFAM01.1	PPB	1/3/2025 12:09	SMPL
P5393-05	MH2D40	276	-0.0042 -		SFAM01.1	PPB	1/3/2025 12:11	SMPL
P5393-06	MH2D43	404	0.0256 -		SFAM01.1	PPB	1/3/2025 12:13	SMPL
P5393-07	MH2D46	290	-0.001 -		SFAM01.1	PPB	1/3/2025 12:16	SMPL
P5393-08	MH2D49	384	0.0209 -		SFAM01.1	PPB	1/3/2025 12:18	SMPL
P5393-09	MH2D52	1144	0.1981 -		SFAM01.1	PPB	1/3/2025 12:20	SMPL
P5393-10	MH2D55	165	-0.0301 -		SFAM01.1	PPB	1/3/2025 12:23	SMPL
P5393-11	MH2D58	810	0.1203 -		SFAM01.1	PPB	1/3/2025 12:25	SMPL
P5393-12	MH2D61	169	-0.0292 -		SFAM01.1	PPB	1/3/2025 12:27	SMPL
P5393-13	MH2D64	134	-0.0374 -		SFAM01.1	PPB	1/3/2025 12:29	SMPL

LBN134154
INSTRUMENT ID : CV1

P5393-14	MH2D67	172	-0.0285 -	SFAM01.1	PPB	1/3/2025 12:32 SMPL
P5393-15	MH2D70	130	-0.0383 -	SFAM01.1	PPB	1/3/2025 12:34 SMPL
P5393-16	MH2D73	262	-0.0075 -	SFAM01.1	PPB	1/3/2025 12:36 SMPL
P5393-17	MH2D76	208	-0.0201 -	SFAM01.1	PPB	1/3/2025 12:38 SMPL
P5393-18	MH2D76D	173	-0.0283 -	SFAM01.1	PPB	1/3/2025 12:41 SMPL
P5393-19	MH2D76S	5423	1.1958 -	SFAM01.1	PPB	1/3/2025 12:43 SMPL
CCV098	CCV098	22453	5.1663 -	SFAM01.1	PPB	1/3/2025 12:45 SMPL
CCB098	CCB098	163	-0.0306 -	SFAM01.1	PPB	1/3/2025 12:47 SMPL
PB165948TB	LEB948	151	-0.0334 -	SFAM01.1	PPB	1/3/2025 12:50 SMPL
PB165949TB	LEB949	166	-0.0299 -	SFAM01.1	PPB	1/3/2025 12:52 SMPL
CCV099	CCV099	22341	5.1402 -	SFAM01.1	PPB	1/3/2025 12:54 SMPL
CCB099	CCB099	176	-0.0276 -	SFAM01.1	PPB	1/3/2025 12:56 SMPL