

MH2D76A

POST-DIGESTION/DISTILLATION SPIKE SAMPLE RECOVERY

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	51811
MA No. :		SDG No.:	MH2D28
Matrix :	Water	Analytical Method:	ICP-AES
% Solids:			
Concentration Units (µg/L, mg/L or mg/kg dry weight):			ug/L

Analyte	Control Limit %R	Spiked Sample Result (SSR) Q		Sample Result (SR) Q		Spike Added (SA)	%R	Q
Arsenic	75 - 125	66		27		54.0	72	
Barium	75 - 125	360		31	J	400	82	