

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134297

Review By	Sarabjit Jaswal	Review On	1/16/2025 1:15:01 PM
Supervise By	Mohan Bera	Supervise On	1/17/2025 1:13:09 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83935,MP84064,MP84062,MP84061,MP84060,MP84059,MP84058		
ICV Standard	MP84065		
CCV Standard	MP84068		
ICSA Standard	MP84066,MP84067		
CRI Standard			
LCS Standard	MP83506		
Chk Standard	MP84071,MP84072		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/15/25 12:32		Kareem	OK
2	S1	S01	CAL2	01/15/25 12:36		Kareem	OK
3	S2	S02	CAL3	01/15/25 12:41		Kareem	OK
4	S3	S03	CAL4	01/15/25 12:45		Kareem	OK
5	S4	S04	CAL5	01/15/25 12:49		Kareem	OK
6	S5	S05	CAL6	01/15/25 12:53		Kareem	OK
7	S6	S06	CAL7	01/15/25 12:57		Kareem	OK
8	ICV004	ICV004	ICV	01/15/25 13:43	Inst. paused for calibration review	Kareem	OK
9	ICB004	ICB004	ICB	01/15/25 13:47		Kareem	OK
10	ICSA004	ICSA004	ICSA	01/15/25 13:51		Kareem	OK
11	ICSAB004	ICSAB004	ICSAB	01/15/25 13:55		Kareem	OK
12	CCV023	CCV023	CCV	01/15/25 14:06	Inst. paused for review	Kareem	OK
13	CCB023	CCB023	CCB	01/15/25 14:11		Kareem	OK
14	PB165925BL	PBS925	MB	01/15/25 14:15		Kareem	OK
15	PB165925BS	LCS925	LCS	01/15/25 14:19		Kareem	OK
16	PB165941BL	PBW941	MB	01/15/25 14:23		Kareem	OK
17	PB165941BS	LCS941	LCS	01/15/25 14:27		Kareem	OK
18	PB165942BL	PBW942	MB	01/15/25 14:31		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134297

Review By	Sarabjit Jaswal	Review On	1/16/2025 1:15:01 PM
Supervise By	Mohan Bera	Supervise On	1/17/2025 1:13:09 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83935,MP84064,MP84062,MP84061,MP84060,MP84059,MP84058		
ICV Standard	MP84065		
CCV Standard	MP84068		
ICSA Standard	MP84066,MP84067		
CRI Standard			
LCS Standard	MP83506		
Chk Standard	MP84071,MP84072		

19	PB165942BS	LCS942	LCS	01/15/25 14:35		Kareem	OK
20	P5388-11DL	MH2D56	SAM	01/15/25 14:39	2x for Fe	Kareem	Confirms
21	P5388-12DL	MH2D59	SAM	01/15/25 14:44	Not Use	Kareem	Not Ok
22	P5393-17	MH2D76	SAM	01/15/25 14:48	Mn High	Kareem	Dilution
23	P5393-18	MH2D76D	DUP	01/15/25 14:52	Mn High	Kareem	Dilution
24	P5393-17L	MH2D76L	SD	01/15/25 14:56	Mn High	Kareem	Dilution
25	P5393-19	MH2D76S	MS	01/15/25 15:00	Mn high, MS fail for As (26.79 ppb)Ba (Below RL).	Kareem	Dilution
26	P5392-17	MH2D75	SAM	01/15/25 15:04	Mn High	Kareem	Dilution
27	P5392-18	MH2D75D	DUP	01/15/25 15:08	Mn High	Kareem	Dilution
28	P5392-17L	MH2D75L	SD	01/15/25 15:13	Mn High	Kareem	Dilution
29	P5392-19	MH2D75S	MS	01/15/25 15:17	Mn high, MS fail for Ba (Below RL).	Kareem	Dilution
30	PB165941TB	LEB941	MB	01/15/25 15:21		Kareem	OK
31	PB165942TB	LEB942	MB	01/15/25 15:25		Kareem	OK
32	P5393-01	MH2D28	SAM	01/15/25 15:29	Mn,Zn High	Kareem	Dilution
33	P5393-02	MH2D31	SAM	01/15/25 15:33	Mn (oversaturated)	Kareem	Dilution
34	CCV024	CCV024	CCV	01/15/25 15:37		Kareem	OK
35	CCB024	CCB024	CCB	01/15/25 15:42		Kareem	OK
36	P5393-03	MH2D34	SAM	01/15/25 15:46	Mn High	Kareem	Dilution
37	P5393-04	MH2D37	SAM	01/15/25 15:50	Cu,Mn,Zn High	Kareem	Dilution

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134297

Review By	Sarabjit Jaswal	Review On	1/16/2025 1:15:01 PM
Supervise By	Mohan Bera	Supervise On	1/17/2025 1:13:09 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83935,MP84064,MP84062,MP84061,MP84060,MP84059,MP84058		
ICV Standard	MP84065		
CCV Standard	MP84068		
ICSA Standard	MP84066,MP84067		
CRI Standard			
LCS Standard	MP83506		
Chk Standard	MP84071,MP84072		

38	P5393-05	MH2D40	SAM	01/15/25 15:54	Ca High	Kareem	Dilution
39	P5393-06	MH2D43	SAM	01/15/25 15:59	Mn High	Kareem	Dilution
40	P5393-07	MH2D46	SAM	01/15/25 16:03		Kareem	OK
41	P5393-08	MH2D49	SAM	01/15/25 16:07	Mn High	Kareem	Dilution
42	P5393-09	MH2D52	SAM	01/15/25 16:11		Kareem	OK
43	P5393-10	MH2D55	SAM	01/15/25 16:15		Kareem	OK
44	P5393-11	MH2D58	SAM	01/15/25 16:19	Cu High	Kareem	Dilution
45	P5393-12	MH2D61	SAM	01/15/25 16:23		Kareem	OK
46	P5393-13	MH2D64	SAM	01/15/25 16:27		Kareem	OK
47	P5393-14	MH2D67	SAM	01/15/25 16:31	Mn,Zn High	Kareem	Dilution
48	P5393-15	MH2D70	SAM	01/15/25 16:35	Mn,Zn High	Kareem	Dilution
49	P5393-16	MH2D73	SAM	01/15/25 16:40	Mn High	Kareem	Dilution
50	P5392-01	MH2D27	SAM	01/15/25 16:44	Mn,Zn High	Kareem	Dilution
51	P5392-02	MH2D30	SAM	01/15/25 16:48	Mn (oversaturated)	Kareem	Dilution
52	P5392-03	MH2D33	SAM	01/15/25 16:52		Kareem	OK
53	P5392-04	MH2D36	SAM	01/15/25 16:56	Cu,Mn,Zn High	Kareem	Dilution
54	P5392-05	MH2D39	SAM	01/15/25 17:01	Ca High	Kareem	Dilution
55	P5392-06	MH2D42	SAM	01/15/25 17:05	Mn High	Kareem	Dilution
56	CCV025	CCV025	CCV	01/15/25 17:25		Kareem	OK
57	CCB025	CCB025	CCB	01/15/25 17:29		Kareem	OK