

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51811 MA No. :                      SDG No.: MH2D28  
 Initial Calibration Verification Source : MP83861  
 Continuing Calibration Verification Source : MP83863  
 Run Batch: LB134154 Analytical Method: CVAA  
 Concentration Units: µg/L

|         | Initial Calibration Verification |       |    |      | Continuing Calibration Verification |       |     |      |            |     |      |
|---------|----------------------------------|-------|----|------|-------------------------------------|-------|-----|------|------------|-----|------|
|         | ID: ICV041                       |       |    |      | ID: CCV096                          |       |     |      | ID: CCV097 |     |      |
| Analyte | True                             | Found | %R | %RSD | True                                | Found | %R  | %RSD | Found      | %R  | %RSD |
| Mercury | 4.0                              | 3.9   | 98 |      | 5.0                                 | 5.0   | 100 |      | 5.1        | 103 |      |

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51811 MA No. :                      SDG No.: MH2D28  
Initial Calibration Verification Source : MP83861  
Continuing Calibration Verification Source : MP83863  
Run Batch: LB134154 Analytical Method: CVAA  
Concentration Units: µg/L

|         | Initial Calibration Verification |       |    |      | Continuing Calibration Verification |       |     |      |            |     |      |
|---------|----------------------------------|-------|----|------|-------------------------------------|-------|-----|------|------------|-----|------|
|         | ID:                              |       |    |      | ID: CCV098                          |       |     |      | ID: CCV099 |     |      |
| Analyte | True                             | Found | %R | %RSD | True                                | Found | %R  | %RSD | Found      | %R  | %RSD |
| Mercury |                                  |       |    |      | 5.0                                 | 5.2   | 103 |      | 5.1        | 103 |      |