



**284 Sheffield Street  
Mountainside, NJ 07092**

## **SDG NARRATIVE**

**USEPA**

**SDG # MH2D28**

**CASE # 51811**

**CONTRACT # 68HERH20D0011**

**SOW# SFAM01.1**

**LAB NAME: Alliance Technical Group, LLC**

**LAB CODE: ACE**

**LAB ORDER ID #P5393**

### **A. Number of Samples and Date of Receipt**

17 Soil samples were delivered to the laboratory intact on 12/27/2024.

### **B. Parameters**

Test requested for SPLP MetalGroup3 = Aluminum, Antimony, Arsenic, Barium, Beryllium, Cadmium, Calcium, Chromium, Cobalt, Copper, Iron, Lead, Magnesium, Manganese, Nickel, Potassium, Selenium, Silver, Sodium, Thallium, Vanadium, Zinc & SPLP Mercury.

### **C. Cooler Temp**

Indicator Bottle: Presence/Absence

Cooler: 3.2°C, 3.1°C, 2.4°C, 3.0°C

### **D. Detail Documentation (related to Sample Handling Shipping, Analytical Problem, Temp of Cooler etc):**

Issue : A "P" or "M" prefix was listed at the beginning of a CLP sample ID.

### **E. Corrective Action taken for above:**

Resolution: To maintain COC integrity, ASB requests no changes to the Sample IDs. The laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples.

### **F. Analytical Techniques:**

All analyses were based on CLP Methodology by method SFAM01.1.



**284 Sheffield Street  
Mountainside, NJ 07092**

Inter Element correction factors (IECs) are determined annually and correction factor are applied during ICP-AES analysis.

#### **G. Calculation:**

##### **Calculation for ICP-AES Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \frac{V_f}{V_i} \times \text{DF} \times 1000$$

Where,

C = Instrument value in ppm (The average of all replicate exposures)

V<sub>f</sub> = Final digestion volume (mL)

V<sub>i</sub> = Initial aliquot amount (mL) (Sample amount taken in prep)

DF = Dilution Factor

##### **Example Calculation For Sample MH2D28 For Arsenic:**

If C = 0.0139316 ppm

V<sub>f</sub> = 50 ml

V<sub>i</sub> = 50 ml

DF = 1

$$\text{Concentration or Result } (\mu\text{g/L}) = 0.0139316 \times \frac{50}{50} \times 1 \times 1000$$

$$= 13.9316 \mu\text{g/L}$$

$$= 14 \mu\text{g/L} \text{ (Reported Result with Signification)}$$

##### **Calculation for Hg Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \text{DF}$$

Where,

C = Instrument response in  $\mu\text{g/L}$  from the calibration curve.

DF = Dilution Factor

##### **Example Calculation For Sample MH2D28:**

If C = 0.0389 ppb

DF = 1



**284 Sheffield Street  
Mountainside, NJ 07092**

Concentration or Result ( $\mu\text{g/L}$ ) =  $0.0389 \times 1$   
=  $0.0389 \mu\text{g/L}$   
=  $0.039 \mu\text{g/L}$  (Reported Result with Signification)

#### **H. QA/ QC**

Calibrations met requirements. Interference check met requirements. Blank analyses did not indicate any presence of contamination. Laboratory Control sample was within control limits. Spike sample did meet requirements except for Antimony, Arsenic, Barium, Silver. Duplicate sample did meet requirements. Serial Dilution did meet requirements.

Samples receive as soil but as per ASR process for SPLP and forms are reported with water.

I certify that the data package is in compliance with the terms and conditions of the contract both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Director or his designee, as verified by the following signature.

Signature \_\_\_\_\_

Name: Nimisha Pandya

Date \_\_\_\_\_

Title: Document Control Officer