

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5394 SAS No.: P5394 SDG NO.: P5394

EPA Sample No.: SSTD020590 Date Analyzed: 12/31/2024

Lab File ID: BG063915.D Time Analyzed: 10:06

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	71358	7.793	283124	10.58	189907	14.44
UPPER LIMIT	142716	8.293	566248	11.084	379814	14.938
LOWER LIMIT	35679	7.293	141562	10.084	94953.5	13.938
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	57403	7.80	215925	10.58	149238	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.