

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5394 SAS No.: P5394 SDG NO.: P5394

EPA Sample No.: SSTD020590 Date Analyzed: 12/31/2024

Lab File ID: BG063915.D Time Analyzed: 10:06

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	447856	17.194	527679	21.448	548626	24.468
UPPER LIMIT	895712	17.694	1055360	21.948	1097250	24.968
LOWER LIMIT	223928	16.694	263840	20.948	274313	23.968
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	358336	17.20	450983	21.45	469321	24.47

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.