

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P5394 SAS No.: P5394 SDG NO.: P5394
 Lab File ID: VN085339.D Date Analyzed: 12/30/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	166322	8.22	293676	9.10	265343	11.87
UPPER LIMIT	332644	8.724	587352	9.6	530686	12.365
LOWER LIMIT	83161	7.724	146838	8.6	132672	11.365
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	135951	8.22	251018	9.10	216231	11.87
STORAGE-BLANK-SOIL-REF-5	134403	8.22	251622	9.10	212421	11.87
STORAGE-BLANK-SOIL-REF-4	140558	8.22	267601	9.10	230526	11.87
STORAGE-BLANK-SAMPLE-MANAGEMENT	130142	8.22	246892	9.10	218306	11.87
VN1230WBL01	161031	8.22	303169	9.10	251914	11.87
VN1230WBS01	172604	8.22	306320	9.10	265813	11.87
VN1230WBSD01	164383	8.22	292155	9.10	254855	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.