

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5395

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL041835.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0103ABS01	Vinyl Chloride	10	8.50	ppbv	85			70	130	
	Heptane	10	8.70	ppbv	87			70	130	
	1,1-Dichloroethene	10	8.20	ppbv	82			70	130	
	Cyclohexane	10	8.30	ppbv	83			70	130	
	cis-1,2-Dichloroethene	10	8.30	ppbv	83			70	130	
	1,1,1-Trichloroethane	10	8.50	ppbv	85			70	130	
	2,2,4-Trimethylpentane	10	10.4	ppbv	104			70	130	
	Benzene	10	9.70	ppbv	97			70	130	
	Trichloroethene	10	8.70	ppbv	87			70	130	
	Toluene	10	9.10	ppbv	91			70	130	
	1,2-Dibromoethane	10	9.70	ppbv	97			70	130	
	Tetrachloroethene	10	8.50	ppbv	85			70	130	
	Ethyl Benzene	10	9.70	ppbv	97			70	130	
	m/p-Xylene	20	20.6	ppbv	103			70	130	
	o-Xylene	10	10.2	ppbv	102			70	130	
	1,3,5-Trimethylbenzene	10	10.7	ppbv	107			70	130	
	1,2,4-Trimethylbenzene	10	11.1	ppbv	111			70	130	
	Naphthalene	10	11.0	ppbv	110			70	130	
	Hexane	10	8.70	ppbv	87			70	130	