

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P5399 SAS No.: P5399 SDG No.: P5399

Instrument ID: MSVOA_N Calibration Date/Time: 01/02/2025 10:36

Lab File ID: VN085372.D Init. Calib. Date(s): 12/20/2024 12/20/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:05 11:42

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.646	1.621	0.5	-1.52	
Toluene	1.843	1.782	0.4	-3.31	
Ethyl Benzene	1.939	1.838	0.1	-5.21	
m/p-Xylenes	0.728	0.731	0.3	0.41	
o-Xylene	0.693	0.673	0.3	-2.89	
1,2-Dichloroethane-d4	2.639	2.614	0.01	-0.95	
Toluene-d8	1.465	1.488	0.01	1.57	
4-Bromofluorobenzene	0.486	0.486	0.2	0	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.