

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5399

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VN085373.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0102WBS01	Benzene	20	18.4	ug/L	92			65	135	
	Toluene	20	18.6	ug/L	93			70	130	
	Ethyl Benzene	20	17.9	ug/L	90			60	140	
	m/p-Xylenes	40	37.2	ug/L	93			87	111	
	o-Xylene	20	17.9	ug/L	90			87	111	