

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: P5399 SAS No.: P5399 SDG NO.: P5399
 Lab File ID: VN085372.D Date Analyzed: 01/02/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:36
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	30931	7.81	161808	9.10	150173	11.87
UPPER LIMIT	61862	8.312	323616	9.6	300346	12.365
LOWER LIMIT	15465.5	7.312	80904	8.6	75086.5	11.365
EPA SAMPLE NO.						
001 WILLETS PT BLVD (DEC)	29248	7.81	147571	9.10	145951	11.87
002 35TH AVE (DEC)	27625	7.81	134721	9.10	130305	11.86
VN0102WBL01	28991	7.81	140575	9.10	123860	11.87
VN0102WBS01	34895	7.81	188370	9.10	177459	11.87
VN0102WBSD01	28825	7.81	150375	9.10	140882	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.