

DATA PACKAGE

SUB - DATA

PROJECT NAME : TRANSFER STATION-SPDES

TULLY ENVIRONMENTAL, INC

127-50 Northern Blvd.

Flushing, NY - 11368

Phone No: 718-446-7000

ORDER ID : P5400

ATTENTION : Dean Devoe



Cover Page

Order ID : P5400

Project ID : Transfer Station-SPDES

Client : Tully Environmental, Inc

Lab Sample Number

P5400-01

Client Sample Number

002 35th Ave (Dec)

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 11:19 am, Jan 14, 2025

Date: 1/10/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

January 07, 2025

PM
AAS-NEW
284 Sheffield Street
Mountainside, NJ 07092
TEL:
FAX:
RE: P5400

Dear PM:

Order No.: 25010114

Summit Environmental Technologies, Inc. received 1 sample(s) on 1/2/2025 for the analyses presented in the following report.

There were no problems with the analytical events associated with this report unless noted in the Case Narrative.

Quality control data is within laboratory defined or method specified acceptance limits except where noted.

If you have any questions regarding these tests results, please feel free to call the laboratory.

Sincerely,

Jennifer Woolf

Project Manager

3310 Win St.
Cuyahoga Falls, Ohio 44223

Arkansas 88-0735, California 2943, Colorado, Connecticut PH-0108, Florida NELAC E87688, Idaho OH00923, Illinois 200061, Indiana C-OH-13, ISO/IEC 17025:2017 119125 L22-544, Kansas E-10347, Kentucky (Underground Storage Tank) 3, Kentucky 90146, Maryland 339, Michigan 9988, Minnesota 1780279, Nevada OH009232020-1, New Hampshire 2996, New Jersey OH006, New York 11777, North Carolina 39705 and 631, North Dakota R-201, Ohio DW, Ohio VAP CL0052, Oklahoma 2019-155, Oregon OH200001, Pennsylvania 68-01335, Rhode Island LA000317, South Carolina 92016001, Texas T104704466-19-16, Utah OH009232020-12, Virginia VELAP 10381, West Virginia 9957C



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Case Narrative

WO#: 25010114
Date: 1/7/2025

CLIENT: AAS-NEW

Project: P5400

WorkOrder Narrative:

25010114: This report in its entirety consists of the following documents: Cover Letter, Case Narrative, Analytical Results, QC Summary Report, Applicable Accreditation Information, Chain-of-Custody, Cooler Receipt Form, and other applicable forms as necessary. All documents contain the Summit Environmental Technologies, Inc., Work Order Number assigned to this report.

Summit Environmental Technologies, Inc., holds the accreditations/certifications listed at the bottom of the cover letter that may or may not pertain to this report. Please refer to the "Accreditation Program Analytes Report" for accredited analytes list.

The information contained in this analytical report is the sole property of Summit Environmental Technologies, Inc. and that of the customer. It cannot be reproduced in any form without the consent of Summit Environmental Technologies, Inc. or the customer for which this report was issued. The results contained in this report are only representative of the samples received. Conditions can vary at different times and at different sampling conditions. Summit Environmental Technologies, Inc. is not responsible for use or interpretation of the data included herein.

All results for Solid Samples are reported on an "as received" or "wet weight" basis unless indicated as "dry weight" using the "-dry" designation on the reporting units.

This report is believed to meet all of the requirements of the accrediting agency, where applicable. Any comments or problems with the analytical events associated with this report are noted below.

Analytical Sequence Sample Notes:

25010114-001A HG-LL_NPW(1631): Z: Method Deviation: Sample was received without an associated Field or Trip Blank for Low Level Mercury Analysis.

Original

These commonly used Qualifiers and Acronyms may or may not be present in this report.

Qualifiers

U	The compound was analyzed for but was not detected above the MDL.
J	The reported value is greater than the Method Detection Limit but less than the Reporting Limit.
H	The hold time for sample preparation and/or analysis was exceeded. Not Clean Water Act compliant.
D	The result is reported from a dilution.
E	The result exceeded the linear range of the calibration or is estimated due to interference.
MC	The result is below the Minimum Compound Limit.
*	The result exceeds the Regulatory Limit or Maximum Contamination Limit.
m	Manual integration was used to determine the area response.
d	Manual integration in which peak was deleted
N	The result is presumptive based on a Mass Spectral library search assuming a 1:1 response.
P	The second column confirmation exceeded 25% difference.
C	The result has been confirmed by GC/MS.
X	The result was not confirmed when GC/MS Analysis was performed.
B	The analyte was detected in the Method Blank at a concentration greater than the RL.
MB+	The analyte was detected in the Method Blank at a concentration greater than the MDL.
G	The ICB or CCB contained reportable amounts of analyte.
QC-/+	The CCV recovery failed low (-) or high (+).
R/QDR	The RPD was outside of accepted recovery limits.
QL-/+	The LCS or LCSD recovery failed low (-) or high (+).
QLR	The LCS/LCSD RPD was outside of accepted recovery limits.
QM-/+	The MS or MSD recovery failed low (-) or high (+).
QMR	The MS/MSD RPD was outside of accepted recovery limits.
QV-/+	The ICV recovery failed low (-) or high (+).
S	The spike result was outside of accepted recovery limits.
W	Samples were received outside temperature limits (0° – 6° C). Not Clean Water Act compliant.
Z	Deviation; A deviation from the method was performed; Please refer to the Case Narrative for additional information

Acronyms

ND	Not Detected	RL	Reporting Limit
QC	Quality Control	MDL	Method Detection Limit
MB	Method Blank	LOD	Level of Detection
LCS	Laboratory Control Sample	LOQ	Level of Quantitation
LCSD	Laboratory Control Sample Duplicate	PQL	Practical Quantitation Limit
QCS	Quality Control Sample	CRQL	Contract Required Quantitation Limit
DUP	Duplicate	PL	Permit Limit
MS	Matrix Spike	RegLvl	Regulatory Limit
MSD	Matrix Spike Duplicate	MCL	Maximum Contamination Limit
RPD	Relative Percent Different	MinCL	Minimum Compound Limit
ICV	Initial Calibration Verification	RA	Reanalysis
ICB	Initial Calibration Blank	RE	Reextraction
CCV	Continuing Calibration Verification	TIC	Tentatively Identified Compound
CCB	Continuing Calibration Blank	RT	Retention Time
RLC	Reporting Limit Check	CF	Calibration Factor

This list of Qualifiers and Acronyms reflects the most commonly utilized Qualifiers and Acronyms for reporting. Please refer to the Analytical Notes in the Case Narrative for any Qualifiers or Acronyms that do not appear in this list or for additional information regarding the use of these Qualifiers on reported data.



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Workorder Sample Summary

WO#: 25010114
07-Jan-25

CLIENT: AAS-NEW
Project: P5400

Lab SampleID	Client Sample ID	Tag No	Date Collected	Date Received	Matrix
25010114-001	002 35th Ave (Dec)		12/30/2024 11:15:00 AM	1/2/2025 12:20:00 PM	Non-Potable Water



SUMMIT
ENVIRONMENTAL TECHNOLOGIES, INC.
Analytical Laboratories

Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

DATES REPORT

WO#: 25010114

07-Jan-25

Client: AAS-NEW

Project: P5400

Sample ID	Client Sample ID	Collection Date	Matrix	Test Name	Leachate Date	Prep Date	Analysis Date
25010114-001A	002 35th Ave (Dec)	12/30/2024 11:15:00 AM	Non-Potable Water	Low-Level Mercury (EPA 1631)			1/7/2025 10:06:25 AM

Original



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25010114
Date Reported: 1/7/2025
Company: AAS-NEW
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 1/2/2025
Project#: P5400

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
002 35th Ave (Dec)	001	12/30/2024	Mercury	36.1	ng/L	Z	Non-Potable Water	EPA 1631 E	1	0.416	0.500	1/7/2025	TAL

NOTES:

Z: Method Deviation: Sample was received without an associated Field or Trip Blank for Low Level Mercury Analysis.



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4486
Website: <http://www.settek.com>

Accreditation Program Analytes Report

WO#: 25010114
07-Jan-25

Client: AAS-NEW

State: NY

Project: P5400

Program Name: DW_WW_SCM_NI

Sample ID	Matrix	Test Name	Analyte	Status
25010114-001A	Non-Potable Water	Low-Level Mercury (EPA 1631)	Mercury	A

AL	U	Unavailable	LA-NELA	A	Accredited	CO	U	Unavailable
CT	A	Accredited	IL-NELA	A	Accredited	HI-DW	U	Unavailable
IA	N	Not Accredited	IL-NELA	A	Accredited	IN_DW	U	Unavailable
S - NELA	N	Not Accredited	KY_UST	N	Not Accredited	W(RADS)	A	Accredited

Original #1



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25010114
07-Jan-25

Client: AAS-NEW
Project: P5400

BatchID: R200916

Sample ID: RLC	SampType: RLC	TestCode: HG-LL_NPW(Units: ng/L				Prep Date:				RunNo: 200916		
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631				Analysis Date: 1/7/2025				SeqNo: 5401831		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
Mercury	0.948	0.500	0.5000	0	190	50	150				S	

Sample ID: LCS	SampType: LCS	TestCode: HG-LL_NPW(Units: ng/L				Prep Date:			RunNo: 200916		
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631				Analysis Date: 1/7/2025			SeqNo: 5401835		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Mercury	49.6	0.500	50.00	0	99.1	72	128				

Sample ID: mblank1	SampType: MBLK	TestCode: HG-LL_NPW(Units: ng/L			Prep Date:				RunNo: 200916			
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631			Analysis Date: 1/7/2025				SeqNo: 5401839			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
Mercury	ND	0.500									U	

Sample ID: mblank2	SampType: MBLK	TestCode: HG-LL_NPW(Units: ng/L				Prep Date:				RunNo: 200916		
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631				Analysis Date: 1/7/2025				SeqNo: 5401857		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
Mercury	ND	0.500									U	

Qualifiers:

H Holding times for preparation or analysis exceeded
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

ND Not Detected
S Spike Recovery outside accepted recovery limits

Original



Summit Environmental Technologies, Inc.
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25010114
07-Jan-25

Client: AAS-NEW
Project: P5400

BatchID: R200916

Sample ID: mblank2		SampType: MBLK		TestCode: HG-LL_NPW(Units: ng/L		Prep Date:				RunNo: 200916			
Client ID: BatchQC		Batch ID: R200916		TestNo: E1631		Analysis Date: 1/7/2025				SeqNo: 5401857			
Analyte		Result		PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Sample ID: LFB	SampType: LCS	TestCode: HG-LL_NPW(Units: ng/L				Prep Date:				RunNo: 200916		
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631				Analysis Date: 1/7/2025				SeqNo: 5401858		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
Mercury	49.4	0.500	50.00	0	98.8	72	128					

Sample ID: LFBD	SampType: LCSD	TestCode: HG-LL_NPW(Units: ng/L				Prep Date:			RunNo: 200916		
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631				Analysis Date: 1/7/2025			SeqNo: 5401859		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Mercury	50.0	0.500	50.00	0	99.9	72	128	49.39	1.16	24	

Sample ID: mblank3	SampType: MBLK	TestCode: HG-LL_NPW(Units: ng/L				Prep Date:				RunNo: 200916		
Client ID: BatchQC	Batch ID: R200916	TestNo: E1631				Analysis Date: 1/7/2025				SeqNo: 5401861		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
Mercury	ND	0.500									U	

Qualifiers: H Holding times for preparation or analysis exceeded
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

ND Not Detected
S Spike Recovery outside accepted recovery limits

Original

P5400



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax (908) 789-8922
WWW.CHEMTECH.NET

25010114

CHAIN OF CUSTODY RECORD

Sub Lab INFORMATION	CLIENT PROJECT INFORMATION	CLIENT BILLING INFORMATION
COMPANY : Summit Enviromental Technologies I	ORDER ID : P5400	BILL TO: CHEMTECH PO# : P5400
ADDRESS : 3310 Win St	PROJECT ID:Transfer Station-SPDES	ADDRESS : 284, Sheffield Street
CITY:Cuyahoga State :Ohio ZIP :44223	PROJECT MANAGER Yazmeen	CITY: Mountainside State : NJ ZIP : 07092
E-mail :	E-mail : YAZMEEN@CHEMTECH.NET	ATTENTION :Yazmeen
PHONE :330-253-8211	PHONE : (908) 789 8900 FAX: (908) 789 8922	PHONE : (908) 789 8900 FAX : (908) 789 8922

EDD : EXCEL NOCLEAN	Report : Results Only	Comment :
---------------------	-----------------------	-----------

ID	CLIENT SAMPLE IDENTIFICATION	SAMPLE MATRIX	ANALYSIS	Preservative	Method	SAMPLE COLLECTION		# OF BOTTLES	TAT DAYS
						DATE	TIME		
01	002 35th Ave (Dec)	Water	Low-Level Mercury - sub	1:1 HCl to pH < 2	1631	12/30/2024	11:15:00	1	5

TRK. 7711 5868 7650

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGES POSSESSION INCLUDING COURIER DELIVERY									
RELINQUISHED BY SAMPLER: 1. JT	DATETIME: 1220 1-2-25	RECEIVED BY: 1. [Signature]	Conditions of bottles or Coolers at receipt: 11/3/25 1140 Fax 5/10 of 110 = 5.1				<input type="checkbox"/> Compliant <input type="checkbox"/> Non Compliant	Cooler Temp _____ Ice or Cooler? _____	
RELINQUISHED BY: 2.	DATETIME:	RECEIVED BY: 2.							
RELINQUISHED BY: P5400	DATETIME:	RECEIVED BY:					<input type="checkbox"/> OVERNIGHT	Shipment Complete: 13 of 14	

Sample Log-In Check List

Client Name: CHE-NJ-07972

Work Order Number: 25010114

RcptNo: 1

Logged by: Spencer M. Hartwell 1/2/2025 12:20:00 PM

Spencer M. Hartwell

Completed By: Tegan A. Richards 1/3/2025 3:20:39 PM

Tegan Richards

Reviewed By: Jennifer Woolf 1/3/2025 3:35:05 PM

Jennifer M. Woolf

Chain of Custody

1. Is Chain of Custody complete? Yes ☒ No ☐ Not Present ☐
2. How was the sample delivered? FedEx
Tracking No.: 771158687650

Log In

3. Coolers are present? Yes ☒ No ☐ NA ☐
4. Shipping container/cooler in good condition? Yes ☒ No ☐
Custody seals intact on shipping container/cooler? Yes ☐ No ☐ Not Present ☒
No. Seal Date: Signed By:
5. Was an attempt made to cool the samples? Yes ☒ No ☐ NA ☐
6. Were all samples received at a temperature of >0° C to 6.0°C Yes ☒ No ☐ NA ☐
7. Sample(s) in proper container(s)? Yes ☒ No ☐
8. Sufficient sample volume for indicated test(s)? Yes ☒ No ☐
9. Are samples (except VOA and ONG) properly preserved? Yes ☒ No ☐
10. Was preservative added to bottles? Yes ☐ No ☒ NA ☐
11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes ☐ No ☐ No VOA Vials ☒
12. Were any sample containers received broken? Yes ☐ No ☒
13. Does paperwork match bottle labels? Yes ☒ No ☐
(Note discrepancies on chain of custody)
14. Are matrices correctly identified on Chain of Custody? Yes ☒ No ☐
15. Is it clear what analyses were requested? Yes ☒ No ☐
16. Were all holding times able to be met? Yes ☒ No ☐
(If no, notify customer for authorization.)

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes ☐ No ☐ NA ☒

Person Notified: Date:
By Whom: Via: ☐ eMail ☐ Phone ☐ Fax ☐ In Person
Regarding:
Client Instructions:

18. Additional remarks:

Cooler Information

Cooler No	Temp °C	Condition	Seal Intact	Seal No	Seal Date	Signed By
1	5.1	Good	Not Present			