



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Summit Environmental Technologies, LLC

SDG No.: P5403

Project: 24121022

RunNo.: LB134135

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TOC	ICV1	mg/L	10.2	10	102	90-110	12/27/2024
Sample ID: TOC	CCV1	mg/L	0.73	10	7	90-110	12/31/2024
Sample ID: TOC	CCV2	mg/L	9.7	10	97	90-110	12/31/2024
Sample ID: TOC	CCV3	mg/L	21.1	10	211	90-110	12/31/2024