

Initial and Continuing Calibration Verification

Client: Summit Environmental Technologies, LLC

SDG No.: P5404

Project: 24120422

RunNo.: LB134141

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	CCV1						
TOC		mg/L	1.5	1	150	50-150	01/03/2025
Sample ID:	CCV2						
TOC		mg/L	10.7	10	107	90-110	01/03/2025
Sample ID:	CCV3						
TOC		mg/L	19.7	20	99	90-110	01/04/2025
Sample ID:	CCV4						
TOC		mg/L	0.98	1	98	50-150	01/04/2025
Sample ID:	CCV5						
TOC		mg/L	10.8	10	108	90-110	01/08/2025
Sample ID:	CCV6						
TOC		mg/L	21.6	20	108	90-110	01/08/2025
Sample ID:	ICV1						
TOC		mg/L	10.2	10	102	90-110	12/27/2024