



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Summit Environmental Technologies, LLC

SDG No.: P5406

Project: 24121564

RunNo.: LB134208

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	CCV1						
TOC		mg/L	10.4	10	104	90-110	01/09/2025
Sample ID:	CCV2						
TOC		mg/L	10.5	10	105	90-110	01/09/2025
Sample ID:	ICV1						
TOC		mg/L	10.2	10	102	90-110	12/27/2024