



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Blank Summary

**Client:** Summit Environmental Technologies, LLC

**SDG No.:** P5407

**Project:** 24121544

**RunNo.:** LB134199

| Analyte                       | Units | Result   | Acceptance Limits | Conc Qual | MDL  | RDL | Analysis Date |
|-------------------------------|-------|----------|-------------------|-----------|------|-----|---------------|
| Sample ID: <b>CCB1</b><br>TOC | mg/L  | 0.25     | 0.5000            | J         | 0.19 | 1   | 01/08/2025    |
| Sample ID: <b>CCB2</b><br>TOC | mg/L  | 0.43     | 0.5000            | J         | 0.19 | 1   | 01/08/2025    |
| Sample ID: <b>CCB3</b><br>TOC | mg/L  | 0.38     | 0.5000            | J         | 0.19 | 1   | 01/09/2025    |
| Sample ID: <b>ICB1</b><br>TOC | mg/L  | < 0.5000 | 0.5000            | U         | 0.19 | 1   | 12/27/2024    |