

Initial and Continuing Calibration Blank Summary

Client: Summit Environmental Technologies, LLC

SDG No.: P5408

Project: 24120596

RunNo.: LB134141

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: CCB1 TOC	mg/L	0.28	0.5000	J	0.19	1	01/03/2025
Sample ID: CCB2 TOC	mg/L	0.4	0.5000	J	0.19	1	01/03/2025
Sample ID: CCB3 TOC	mg/L	0.35	0.5000	J	0.19	1	01/04/2025
Sample ID: CCB4 TOC	mg/L	0.28	0.5000	J	0.19	1	01/04/2025
Sample ID: CCB5 TOC	mg/L	0.31	0.5000	J	0.19	1	01/08/2025
Sample ID: CCB6 TOC	mg/L	0.35	0.5000	J	0.19	1	01/08/2025
Sample ID: ICB1 TOC	mg/L	< 0.5000	0.5000	U	0.19	1	12/27/2024