

Instrument ID: TOC

Daily Analysis Runlog For Sequence/QC Batch ID # LB134199

Review By	Niha	Review On	1/10/2025 10:31:09 AM
Supervise By	Iwona	Supervise On	1/10/2025 10:32:07 AM
SubDirectory	LB134199	Test	TOC
STD. NAME	STD REF.#		
ICAL Standard	WP111255,WP111256,WP111257,WP111258,WP111259,WP111260		
ICV Standard	WP111262		
CCV Standard	WP111313		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111314		
Chk Standard	WP109953,WP111267,WP111268		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPM	0.0PPM	CAL1	12/27/24 12:55		NF IZ	OK
2	0.5PPM	0.5PPM	CAL2	12/27/24 13:20		NF IZ	OK
3	1.0PPM	1.0PPM	CAL3	12/27/24 14:34		NF IZ	OK
4	2.0PPM	2.0PPM	CAL4	12/27/24 15:01		NF IZ	OK
5	5.0PPM	5.0PPM	CAL5	12/27/24 15:58		NF IZ	OK
6	10.0PPM	10.0PPM	CAL6	12/27/24 16:28		NF IZ	OK
7	20.0PPM	20.0PPM	CAL7	12/27/24 16:59		NF IZ	OK
8	ICV1	ICV1	ICV	12/27/24 17:29		NF IZ	OK
9	ICB1	ICB1	ICB	12/27/24 17:53		NF IZ	OK
10	IC-20	IC-20	SAM	12/27/24 18:40		NF IZ	OK
11	IC-R	IC-R	SAM	12/27/24 19:04		NF IZ	OK
12	CCV1	CCV1	CCV	01/08/25 15:56		NF IZ	OK
13	CCB1	CCB1	CCB	01/08/25 16:21		NF IZ	OK
14	LB134199BLW	LB134199BLW	MB	01/08/25 16:44		NF IZ	OK
15	LB134199BSW	LB134199BSW	LCS	01/08/25 17:10		NF IZ	OK
16	P5407-01	2412-0430-001	SAM	01/08/25 17:38		NF IZ	OK
17	P5407-01MS	2412-0430-001MS	MS	01/08/25 18:06	2.0ml WP111311+38.0ml Sample	NF IZ	OK
18	P5407-01MSD	2412-0430-001MSD	MSD	01/08/25 18:35	2.0ml WP111311+38.0ml Sample	NF IZ	OK

Instrument ID: TOC

Daily Analysis Runlog For Sequence/QC Batch ID # LB134199

Review By	Niha	Review On	1/10/2025 10:31:09 AM
Supervise By	Iwona	Supervise On	1/10/2025 10:32:07 AM
SubDirectory	LB134199	Test	TOC
STD. NAME	STD REF.#		
ICAL Standard	WP111255,WP111256,WP111257,WP111258,WP111259,WP111260		
ICV Standard	WP111262		
CCV Standard	WP111313		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111314		
Chk Standard	WP109953,WP111267,WP111268		

19	P5407-02	2412-0430-002	SAM	01/08/25 19:01		NF IZ	OK
20	P5407-03	2412-0430-003	SAM	01/08/25 20:18		NF IZ	OK
21	P5407-04	2412-0430-004	SAM	01/08/25 20:43		NF IZ	OK
22	P5407-05	2412-0430-005	SAM	01/08/25 21:09		NF IZ	OK
23	P5407-06	2412-0430-006	SAM	01/08/25 21:35		NF IZ	OK
24	CCV2	CCV2	CCV	01/08/25 22:25		NF IZ	OK
25	CCB2	CCB2	CCB	01/08/25 22:48		NF IZ	OK
26	P5407-07	2412-0430-007	SAM	01/09/25 00:03	High	NF IZ	Dilution
27	P5410-01	OUTFALL-003	SAM	01/09/25 00:29		NF IZ	OK
28	P5410-02	OUTFALL-005	SAM	01/09/25 00:54		NF IZ	OK
29	P5410-03	OUTFALL-007	SAM	01/09/25 01:20		NF IZ	OK
30	P5410-04	INFLOW-007	SAM	01/09/25 01:45		NF IZ	OK
31	P5407-07DL	2412-0430-007DL	SAM	01/09/25 09:49	Report 2X	NF IZ	Confirms
32	CCV3	CCV3	CCV	01/09/25 10:15		NF IZ	OK
33	CCB3	CCB3	CCB	01/09/25 10:37		NF IZ	OK