

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134165

Review By	Ketan	Review On	1/6/2025 10:59:55 PM
Supervise By	mohan	Supervise On	1/6/2025 11:01:08 PM

STD. NAME	STD REF.#
ICAL Standard	MP83855,MP83856,MP83857,MP83858,MP83859,MP83860
ICV Standard	MP83861
CCV Standard	MP83863
ICSA Standard	
CRI Standard	MP83865
LCS Standard	
Chk Standard	MP83862,MP83864,MP83866,MP83870

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/06/25 11:45		Mohan	OK
2	S0.2	S0.2	CAL2	01/06/25 11:47		Mohan	OK
3	S2.5	S2.5	CAL3	01/06/25 11:49		Mohan	OK
4	S5	S5	CAL4	01/06/25 11:52		Mohan	OK
5	S7.5	S7.5	CAL5	01/06/25 11:57		Mohan	OK
6	S10	S10	CAL6	01/06/25 12:09		Mohan	OK
7	ICV43	ICV43	ICV	01/06/25 12:17		Mohan	OK
8	ICB43	ICB43	ICB	01/06/25 12:20		Mohan	OK
9	CCV03	CCV03	CCV	01/06/25 12:30		Mohan	OK
10	CCB03	CCB03	CCB	01/06/25 12:32		Mohan	OK
11	CRA	CRA	CRDL	01/06/25 12:37		Mohan	OK
12	HighStd	HighStd	HIGH STD	01/06/25 12:39		Mohan	OK
13	ChkStd	ChkStd	SAM	01/06/25 12:42		Mohan	OK
14	PB165950BL	PB165950BL	MB	01/06/25 12:44		Mohan	OK
15	PB165950BS	PB165950BS	LCS	01/06/25 12:46		Mohan	OK
16	P5398-02	341-349	SAM	01/06/25 12:49		Mohan	OK
17	P5398-02DUP	341-349DUP	DUP	01/06/25 12:53		Mohan	OK
18	P5398-02MS	341-349MS	MS	01/06/25 12:56		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134165

Review By	Ketan	Review On	1/6/2025 10:59:55 PM
Supervise By	mohan	Supervise On	1/6/2025 11:01:08 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83855,MP83856,MP83857,MP83858,MP83859,MP83860		
ICV Standard	MP83861		
CCV Standard	MP83863		
ICSA Standard			
CRI Standard	MP83865		
LCS Standard			
Chk Standard	MP83862,MP83864,MP83866,MP83870		

19	P5398-02MSD	341-349MSD	MSD	01/06/25 12:58		Mohan	OK
20	PB165951BL	PB165951BL	MB	01/06/25 13:00		Mohan	OK
21	CCV04	CCV04	CCV	01/06/25 13:03		Mohan	OK
22	CCB04	CCB04	CCB	01/06/25 13:05		Mohan	OK
23	PB165951BS	PB165951BS	LCS	01/06/25 13:07		Mohan	OK
24	P5386-05	MOO-24-00397	SAM	01/06/25 13:09		Mohan	OK
25	P5386-05DUP	MOO-24-00397DUP	DUP	01/06/25 13:12		Mohan	OK
26	P5386-05MS	MOO-24-00397MS	MS	01/06/25 13:14		Mohan	OK
27	P5386-05MSD	MOO-24-00397MSD	MSD	01/06/25 13:16		Mohan	OK
28	P5411-02	COMP	SAM	01/06/25 13:19		Mohan	OK
29	P5398-02L	341-349L	SD	01/06/25 13:27		Mohan	OK
30	P5398-02A	341-349A	PS	01/06/25 13:29		Mohan	OK
31	P5386-05L	MOO-24-00397L	SD	01/06/25 13:31		Mohan	OK
32	P5386-05A	MOO-24-00397A	PS	01/06/25 13:33		Mohan	OK
33	CCV05	CCV05	CCV	01/06/25 13:36		Mohan	OK
34	CCB05	CCB05	CCB	01/06/25 13:38		Mohan	OK
35	PB165934TB	PB165934TB	MB	01/06/25 13:40		Mohan	OK
36	CCV06	CCV06	CCV	01/06/25 13:42		Mohan	OK
37	CCB06	CCB06	CCB	01/06/25 13:45		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134257

Review By	kareem	Review On	1/14/2025 1:01:51 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	mp83935 mp83936 mp83937 mp83938 mp83939 mp83941		
ICV Standard	mp83942		
CCV Standard	mp83945		
ICSA Standard	mp83943 mp83944		
CRI Standard	mp83941		
LCS Standard			
Chk Standard	mp83948 mp83949		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/13/25 15:53		Kareem	OK
2	S1	S1	CAL2	01/13/25 15:58		Kareem	OK
3	S2	S2	CAL3	01/13/25 16:02		Kareem	OK
4	S3	S3	CAL4	01/13/25 16:06		Kareem	OK
5	S4	S4	CAL5	01/13/25 16:11		Kareem	OK
6	S5	S5	CAL6	01/13/25 16:15		Kareem	OK
7	ICV01	ICV01	ICV	01/13/25 16:19		Kareem	OK
8	LLICV01	LLICV01	LLICV	01/13/25 16:29		Kareem	OK
9	ICB01	ICB01	ICB	01/13/25 16:33		Kareem	OK
10	CRI01	CRI01	CRDL	01/13/25 16:38		Kareem	OK
11	ICSA01	ICSA01	ICSA	01/13/25 16:42		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	01/13/25 16:46		Kareem	OK
13	CCV01	CCV01	CCV	01/13/25 16:51		Kareem	OK
14	CCB01	CCB01	CCB	01/13/25 17:04		Kareem	OK
15	Q1018-01	3732	SAM	01/13/25 17:08		Kareem	OK
16	Q1033-01	28612	SAM	01/13/25 17:13		Kareem	OK
17	Q1018-02	3735	SAM	01/13/25 17:17		Kareem	OK
18	Q1018-03	3744	SAM	01/13/25 17:22		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134257

Review By	kareem	Review On	1/14/2025 1:01:51 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	mp83935 mp83936 mp83937 mp83938 mp83939 mp83941		
ICV Standard	mp83942		
CCV Standard	mp83945		
ICSA Standard	mp83943 mp83944		
CRI Standard	mp83941		
LCS Standard			
Chk Standard	mp83948 mp83949		

19	Q1018-04	3745	SAM	01/13/25 17:26		Kareem	OK
20	Q1021-01	424	SAM	01/13/25 17:31		Kareem	OK
21	CCV02	CCV02	CCV	01/13/25 17:35		Kareem	OK
22	CCB02	CCB02	CCB	01/13/25 17:39		Kareem	OK
23	PB165996BL	PB165996BL	MB	01/13/25 17:44		Kareem	OK
24	PB165996BS	PB165996BS	LCS	01/13/25 17:48	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
25	PB165981BL	PB165981BL	MB	01/13/25 17:52		Kareem	OK
26	PB165981BS	PB165981BS	LCS	01/13/25 17:56	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
27	PB165973TB	PB165973TB	MB	01/13/25 18:07	Sodium,	Kareem	ReRun
28	Q1018-01DUP	3732DUP	DUP	01/13/25 18:11		Kareem	OK
29	Q1018-01L	3732L	SD	01/13/25 18:16		Kareem	OK
30	Q1018-01MS	3732MS	MS	01/13/25 18:20	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
31	CCV03	CCV03	CCV	01/13/25 18:33		Kareem	OK
32	CCB03	CCB03	CCB	01/13/25 18:37		Kareem	OK
33	Q1018-01MSD	3732MSD	MSD	01/13/25 18:41	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
34	Q1018-01A	3732A	PS	01/13/25 18:45	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134257

Review By	kareem	Review On	1/14/2025 1:01:51 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	mp83935 mp83936 mp83937 mp83938 mp83939 mp83941		
ICV Standard	mp83942		
CCV Standard	mp83945		
ICSA Standard	mp83943 mp83944		
CRI Standard	mp83941		
LCS Standard			
Chk Standard	mp83948 mp83949		

35	P5411-02	COMP	SAM	01/13/25 18:50		Kareem	OK
36	Q1032-01DL	50492DL	SAM	01/13/25 18:54	Not required	Kareem	Not Ok
37	Q1033-01DUP	28612DUP	DUP	01/13/25 18:58		Kareem	OK
38	Q1033-01L	28612L	SD	01/13/25 19:03		Kareem	OK
39	Q1033-01MS	28612MS	MS	01/13/25 19:07	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
40	Q1033-01MSD	28612MSD	MSD	01/13/25 19:11	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
41	Q1033-01A	28612A	PS	01/13/25 19:15	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
42	CCV04	CCV04	CCV	01/13/25 19:20		Kareem	OK
43	CCB04	CCB04	CCB	01/13/25 19:24		Kareem	OK
44	LR1	LR1	HIGH STD	01/13/25 19:28		Kareem	OK
45	LR2	LR2	HIGH STD	01/13/25 19:33		Kareem	OK
46	CCV05	CCV05	CCV	01/13/25 19:37		Kareem	OK
47	CCB05	CCB05	CCB	01/13/25 19:41		Kareem	OK
48	Q1032-01	50492	SAM	01/13/25 19:45		Kareem	OK
49	CCV06	CCV06	CCV	01/13/25 19:50		Kareem	OK
50	CCB06	CCB06	CCB	01/13/25 19:54		Kareem	OK