

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P5411
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV43	Mercury	4.00	4.0	100	90 - 110	CV	01/06/2025	12:17	LB134165

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P5411
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Mercury	5.48	5.0	110	90 - 110	CV	01/06/2025	12:30	LB134165
CCV04	Mercury	5.48	5.0	110	90 - 110	CV	01/06/2025	13:03	LB134165
CCV05	Mercury	4.59	5.0	92	90 - 110	CV	01/06/2025	13:36	LB134165
CCV06	Mercury	4.59	5.0	92	90 - 110	CV	01/06/2025	13:42	LB134165

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P5411
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	490	510	96	90 - 110	P	01/13/2025	16:19	LB134257
	Copper	511	510	100	90 - 110	P	01/13/2025	16:19	LB134257
	Lead	978	1000	98	90 - 110	P	01/13/2025	16:19	LB134257
	Nickel	498	530	94	90 - 110	P	01/13/2025	16:19	LB134257
	Zinc	1050	1000	105	90 - 110	P	01/13/2025	16:19	LB134257

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P5411
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.72	6.0	95	80 - 120	P	01/13/2025	16:29	LB134257
	Copper	21.5	20.0	108	80 - 120	P	01/13/2025	16:29	LB134257
	Lead	11.3	12.0	94	80 - 120	P	01/13/2025	16:29	LB134257
	Nickel	39.0	40.0	98	80 - 120	P	01/13/2025	16:29	LB134257
	Zinc	40.5	40.0	101	80 - 120	P	01/13/2025	16:29	LB134257

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P5411

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2370	2500	95	90 - 110	P	01/13/2025	16:51	LB134257
	Copper	1180	1250	95	90 - 110	P	01/13/2025	16:51	LB134257
	Lead	4730	5000	95	90 - 110	P	01/13/2025	16:51	LB134257
	Nickel	2370	2500	95	90 - 110	P	01/13/2025	16:51	LB134257
	Zinc	2380	2500	95	90 - 110	P	01/13/2025	16:51	LB134257
CCV02	Cadmium	2390	2500	96	90 - 110	P	01/13/2025	17:35	LB134257
	Copper	1200	1250	96	90 - 110	P	01/13/2025	17:35	LB134257
	Lead	4770	5000	95	90 - 110	P	01/13/2025	17:35	LB134257
	Nickel	2390	2500	95	90 - 110	P	01/13/2025	17:35	LB134257
	Zinc	2380	2500	95	90 - 110	P	01/13/2025	17:35	LB134257
CCV03	Cadmium	2390	2500	96	90 - 110	P	01/13/2025	18:33	LB134257
	Copper	1200	1250	96	90 - 110	P	01/13/2025	18:33	LB134257
	Lead	4760	5000	95	90 - 110	P	01/13/2025	18:33	LB134257
	Nickel	2390	2500	95	90 - 110	P	01/13/2025	18:33	LB134257
	Zinc	2490	2500	100	90 - 110	P	01/13/2025	18:33	LB134257
CCV04	Cadmium	2420	2500	97	90 - 110	P	01/13/2025	19:20	LB134257
	Copper	1200	1250	96	90 - 110	P	01/13/2025	19:20	LB134257
	Lead	4830	5000	97	90 - 110	P	01/13/2025	19:20	LB134257
	Nickel	2420	2500	97	90 - 110	P	01/13/2025	19:20	LB134257
	Zinc	2440	2500	98	90 - 110	P	01/13/2025	19:20	LB134257
CCV05	Cadmium	2410	2500	96	90 - 110	P	01/13/2025	19:37	LB134257
	Copper	1200	1250	96	90 - 110	P	01/13/2025	19:37	LB134257
	Lead	4810	5000	96	90 - 110	P	01/13/2025	19:37	LB134257
	Nickel	2400	2500	96	90 - 110	P	01/13/2025	19:37	LB134257
	Zinc	2410	2500	96	90 - 110	P	01/13/2025	19:37	LB134257
CCV06	Cadmium	2440	2500	98	90 - 110	P	01/13/2025	19:50	LB134257
	Copper	1210	1250	97	90 - 110	P	01/13/2025	19:50	LB134257
	Lead	4850	5000	97	90 - 110	P	01/13/2025	19:50	LB134257
	Nickel	2430	2500	97	90 - 110	P	01/13/2025	19:50	LB134257
	Zinc	2480	2500	99	90 - 110	P	01/13/2025	19:50	LB134257