



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** P5411
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	106	40 - 160	CV	01/06/2025	12:37	LB134165
CRI01	Cadmium	5.76	6.0	96	40 - 160	P	01/13/2025	16:38	LB134257
	Copper	21.7	20.0	108	40 - 160	P	01/13/2025	16:38	LB134257
	Lead	10.8	12.0	90	40 - 160	P	01/13/2025	16:38	LB134257
	Nickel	38.7	40.0	97	40 - 160	P	01/13/2025	16:38	LB134257
	Zinc	42.0	40.0	105	40 - 160	P	01/13/2025	16:38	LB134257