

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** P5411
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5411 **SAS No.:** P5411
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	0.85	1.0	85	-5	7	01/13/2025	16:42	LB134257
	Copper	5.77	2.0	288	-18	22	01/13/2025	16:42	LB134257
	Lead	8.17			-12	12	01/13/2025	16:42	LB134257
	Nickel	2.06	2.0	103	-38	42	01/13/2025	16:42	LB134257
	Zinc	5.63			-40	40	01/13/2025	16:42	LB134257
ICSAB01	Cadmium	948	972	98	826	1120	01/13/2025	16:46	LB134257
	Copper	459	511	90	434	588	01/13/2025	16:46	LB134257
	Lead	52.6	49.0	107	37	61	01/13/2025	16:46	LB134257
	Nickel	945	954	99	810	1100	01/13/2025	16:46	LB134257
	Zinc	1030	952	108	809	1095	01/13/2025	16:46	LB134257