

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>P5411</u>
contract: <u>ARAM01</u>	lab code: <u>CHEM</u>	case no.: <u>P5411</u> sas no.: <u>P5411</u>
matrix: <u>Water</u>	sample id: <u>P5386-05</u>	client id: <u>MOO-24-00397MS</u>
Percent Solids for Sample: NA	Spiked ID: P5386-05MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.76		0.18	J	4.0	90		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Aramark Uniforms</u>	level:	<u>low</u>	sdg no.:	<u>P5411</u>	
contract:	<u>ARAM01</u>	lab code:	<u>CHEM</u>	case no.:	<u>P5411</u>	sas no.: <u>P5411</u>
matrix:	<u>Water</u>	sample id:	<u>P5386-05</u>	client id:	<u>MOO-24-00397MSD</u>	
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>P5386-05MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>	

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	4.10		0.18	J	4.0	98		CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>P5411</u>
contract: <u>ARAM01</u>	lab code: <u>CHEM</u>	case no.: <u>P5411</u> sas no.: <u>P5411</u>
matrix: <u>Water</u>	sample id: <u>Q1033-01</u>	client id: <u>28612MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1033-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	93.2		3.00	U	100	93		P
Copper	ug/L	75 - 125	142		10.0	U	150	95		P
Lead	ug/L	75 - 125	449		6.00	U	500	90		P
Nickel	ug/L	75 - 125	236		2.14	J	250	94		P
Zinc	ug/L	75 - 125	124		19.1	J	100	105		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>P5411</u>
contract: <u>ARAM01</u>	lab code: <u>CHEM</u>	case no.: <u>P5411</u> sas no.: <u>P5411</u>
matrix: <u>Water</u>	sample id: <u>Q1033-01</u>	client id: <u>28612MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1033-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	92.7		3.00	U	100	93		P
Copper	ug/L	75 - 125	142		10.0	U	150	95		P
Lead	ug/L	75 - 125	446		6.00	U	500	89		P
Nickel	ug/L	75 - 125	235		2.14	J	250	93		P
Zinc	ug/L	75 - 125	123		19.1	J	100	104		P