

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN010225.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 02 15:39:17 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN035871.D 0.2 =BN035872.D 0.4 =BN035873.D 0.8 =BN035874.D 1.6 =BN035875.D 3.2 =BN035876.D 5.0 =BN035877.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.454	0.422	0.373	0.388	0.391	0.377	0.377	0.397	7.52
3)	n-Nitrosodimet...	0.707	0.674	0.676	0.690	0.722	0.690	0.692	0.693	2.45
4) S	2-Fluorophenol	1.031	1.009	0.952	0.958	0.997	0.956	0.968	0.981	3.13
5) S	Phenol-d6	1.351	1.255	1.180	1.197	1.215	1.163	1.170	1.219	5.44
6)	bis(2-Chloroet...	1.001	0.946	0.936	0.913	0.938	0.886	0.879	0.929	4.43
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.346	0.307	0.296	0.302	0.319	0.317	0.330	0.317	5.48
9)	Naphthalene	1.163	1.094	1.086	1.096	1.167	1.113	1.141	1.123	3.00
10)	Hexachlorobuta...	0.368	0.354	0.353	0.363	0.382	0.362	0.369	0.365	2.74
11) SURR2	Methylnaphth...	0.547	0.536	0.527	0.519	0.556	0.524	0.540	0.536	2.50
12)	2-Methylnaphth...	0.691	0.654	0.685	0.680	0.731	0.701	0.722	0.695	3.75
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.164	0.165	0.189	0.189	0.207	0.211	0.220	0.192	11.39
15) S	2-Fluorobiphenyl	1.776	1.675	1.708	1.765	1.823	1.779	1.762	1.755	2.79
16)	Acenaphthylene	1.890	1.766	1.819	1.839	1.962	1.948	1.963	1.884	4.15
17)	Acenaphthene	1.187	1.162	1.198	1.232	1.300	1.275	1.291	1.235	4.43
18)	Fluorene	1.341	1.270	1.307	1.298	1.419	1.444	1.432	1.359	5.28
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.058	0.066	0.067	0.069	0.076	0.076	0.074	0.069	9.51
21)	4-Bromophenyl-...	0.265	0.269	0.268	0.267	0.290	0.283	0.279	0.274	3.47
22)	Hexachlorobenzene	0.393	0.369	0.356	0.362	0.394	0.373	0.371	0.374	3.93
23)	Atrazine	0.169	0.191	0.176	0.171	0.198	0.190	0.193	0.184	6.36
24)	Pentachlorophenol	0.141	0.101	0.118	0.122	0.143	0.148	0.153	0.132	14.40
25)	Phenanthrene	1.131	1.132	1.142	1.144	1.228	1.193	1.200	1.167	3.36
26)	Anthracene	0.996	0.998	1.008	1.033	1.137	1.132	1.130	1.062	6.34
27) SURR	Fluoranthene-d10	0.988	0.952	0.978	0.959	1.028	1.010	1.035	0.993	3.28
28)	Fluoranthene	1.268	1.253	1.330	1.312	1.441	1.446	1.475	1.361	6.71
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.606	1.620	1.571	1.612	1.711	1.621	1.627	1.624	2.63
31) S	Terphenyl-d14	0.814	0.813	0.790	0.777	0.828	0.776	0.781	0.797	2.64
32)	Benzo(a)anthra...	1.379	1.382	1.344	1.407	1.461	1.427	1.466	1.410	3.19
33)	Chrysene	1.484	1.458	1.441	1.451	1.541	1.478	1.471	1.475	2.24
34)	Bis(2-ethylhex...	0.691	0.583	0.582	0.541	0.569	0.519	0.530	0.574	10.05
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN010225.M

36)	Indeno(1,2,3-c...	1.428	1.461	1.414	1.578	1.697	1.712	1.751	1.577	9.15
37)	Benzo(b)fluora...	1.337	1.304	1.294	1.351	1.466	1.420	1.447	1.374	5.06
38)	Benzo(k)fluora...	1.279	1.254	1.251	1.344	1.469	1.442	1.482	1.360	7.55
39) C	Benzo(a)pyrene	1.099	1.181	1.086	1.163	1.270	1.244	1.284	1.190	6.71
40)	Dibenzo(a,h)an...	1.145	1.152	1.106	1.251	1.363	1.365	1.402	1.255	9.75
41)	Benzo(g,h,i)pe...	1.335	1.316	1.245	1.407	1.490	1.501	1.530	1.403	7.72

(#) = Out of Range