

Surrogate Summary

SW-846

SDG No.: Q1004

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB165947BL	PB165947BL	2-Methylnaphthalene-d10	0.4	0.40	99		30	150
		Fluoranthene-d10	0.4	0.41	101		30	150
		Nitrobenzene-d5	0.4	0.41	103		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.43	107		58	132
PB165947BS	PB165947BS	2-Methylnaphthalene-d10	0.4	0.39	98		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.40	101		55	111
		2-Fluorobiphenyl	0.4	0.39	98		53	106
		Terphenyl-d14	0.4	0.42	105		58	132
PB165947BSD	PB165947BSD	2-Methylnaphthalene-d10	0.4	0.45	113		30	150
		Fluoranthene-d10	0.4	0.45	112		30	150
		Nitrobenzene-d5	0.4	0.45	113	*	55	111
		2-Fluorobiphenyl	0.4	0.46	114	*	53	106
		Terphenyl-d14	0.4	0.50	125		58	132
Q1004-02	VPB190A-HYD-20241231	2-Methylnaphthalene-d10	0.4	0.15	37		30	150
		Fluoranthene-d10	0.4	0.26	65		30	150
		Nitrobenzene-d5	0.4	0.33	82		55	111
		2-Fluorobiphenyl	0.4	0.32	81		53	106
		Terphenyl-d14	0.4	0.40	101		58	132
Q1004-05	BP-VPB-190A-GW-803-805	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.32	81		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.25	63		53	106
		Terphenyl-d14	0.4	0.40	99		58	132