

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0106WBSD01		SDG No.:	Q1004
Lab Sample ID:	VX0106WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044587.D	1		01/06/25 12:00	VX010625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	18.3		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.9		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.0		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	16.4		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.1		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	20.8		0.26	0.75	1.00	ug/L
67-64-1	Acetone	78.5		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	20.8		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.1		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	20.3		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.3		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.9		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	86.4		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.8		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.9		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	19.0		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.7		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	23.0		0.19	0.50	1.00	ug/L
71-43-2	Benzene	22.2		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.3		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	23.0		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.5		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	22.5		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	97.7		0.75	2.50	5.00	ug/L
108-88-3	Toluene	22.3		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	21.6		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	22.9		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	23.2		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	98.0		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0106WBSD01		SDG No.:	Q1004
Lab Sample ID:	VX0106WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044587.D	1		01/06/25 12:00	VX010625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	23.2		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	25.0		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	22.8		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	21.6		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	45.0		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	22.4		0.14	0.50	1.00	ug/L
100-42-5	Styrene	22.7		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	25.0		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	21.1		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.1		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	22.9		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	22.7		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	22.6		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	46.7		81 - 118		93%	SPK: 50
1868-53-7	Dibromofluoromethane	53.2		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	53.9		89 - 112		108%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.2		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	135000	5.544				
540-36-3	1,4-Difluorobenzene	228000	6.757				
3114-55-4	Chlorobenzene-d5	199000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	96100	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products