

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1004

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX044586.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0106WBS01	Chloromethane	20	17.8	ug/L	89			50	139	
	Vinyl chloride	20	17.3	ug/L	86			58	137	
	Bromomethane	20	17.2	ug/L	86			53	141	
	Chloroethane	20	18.2	ug/L	91			60	138	
	Trichlorofluoromethane	20	16.5	ug/L	83			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.2	ug/L	101			70	136	
	1,1-Dichloroethene	20	20.0	ug/L	100			71	131	
	Acetone	100	73.3	ug/L	73			39	160	
	Carbon disulfide	20	20.3	ug/L	102			64	133	
	Methyl tert-butyl Ether	20	17.8	ug/L	89			71	124	
	Methylene Chloride	20	19.7	ug/L	99			74	124	
	trans-1,2-Dichloroethene	20	19.3	ug/L	97			75	124	
	1,1-Dichloroethane	20	18.2	ug/L	91			77	125	
	2-Butanone	100	78.1	ug/L	78			56	143	
	Carbon Tetrachloride	20	22.0	ug/L	110			72	136	
	cis-1,2-Dichloroethene	20	19.1	ug/L	96			78	123	
	Chloroform	20	18.0	ug/L	90			79	124	
	1,1,1-Trichloroethane	20	19.3	ug/L	97			74	131	
	Methylcyclohexane	20	22.5	ug/L	113			72	132	
	Benzene	20	21.5	ug/L	108			79	120	
	1,2-Dichloroethane	20	20.7	ug/L	104			73	128	
	Trichloroethene	20	23.5	ug/L	117			79	123	
	1,2-Dichloropropane	20	21.2	ug/L	106			78	122	
	Bromodichloromethane	20	22.5	ug/L	113			79	125	
	4-Methyl-2-Pentanone	100	93.6	ug/L	94			67	130	
	Toluene	20	22.1	ug/L	111			80	121	
	t-1,3-Dichloropropene	20	21.5	ug/L	108			73	127	
	cis-1,3-Dichloropropene	20	22.6	ug/L	113			75	124	
	1,1,2-Trichloroethane	20	22.2	ug/L	111			80	119	
	2-Hexanone	100	92.7	ug/L	93			57	139	
	Dibromochloromethane	20	22.5	ug/L	113			74	126	
	Tetrachloroethene	20	24.5	ug/L	123			74	129	
	Chlorobenzene	20	22.1	ug/L	111			82	118	
	Ethyl Benzene	20	21.4	ug/L	107			79	121	
	m/p-Xylenes	40	45.1	ug/L	113			80	121	
	o-Xylene	20	22.1	ug/L	111			78	122	
	Styrene	20	22.0	ug/L	110			78	123	
	Bromoform	20	23.2	ug/L	116			66	130	
	Isopropylbenzene	20	21.0	ug/L	105			72	131	
	1,1,2,2-Tetrachloroethane	20	19.2	ug/L	96			71	121	
	1,3-Dichlorobenzene	20	22.1	ug/L	111			80	119	
	1,4-Dichlorobenzene	20	21.7	ug/L	109			79	118	
	1,2-Dichlorobenzene	20	21.2	ug/L	106			80	119	