

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1004 SAS No.: Q1004 SDG NO.: Q1004
Lab File ID: VX044583.D Date Analyzed: 01/06/2025
Instrument ID: MSVOA_X Time Analyzed: 10:04
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	96415	12.018				
UPPER LIMIT	192830	12.518				
LOWER LIMIT	48207.5	11.518				
EPA SAMPLE NO.						
BP-VPB-190A-TB-20241230	104847	12.02				
VPB190A-HYD-20241231	88101	12.02				
BP-VPB-190A-EB-20241231	86329	12.02				
BP-VPB-190A-GW-803-805	86406	12.02				
BP-VPB-190A-GW-818-820	84064	12.02				
VX0106WBL01	101513	12.02				
VX0106WBS01	101841	12.02				
VX0106WBSD01	96127	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.