

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1004 SAS No.: Q1004 SDG NO.: Q1004
Lab File ID: VX044599.D Date Analyzed: 01/07/2025
Instrument ID: MSVOA_X Time Analyzed: 11:14
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	121154	12.018				
UPPER LIMIT	242308	12.518				
LOWER LIMIT	60577	11.518				
EPA SAMPLE NO.						
BP-VPB-190A-GW-778-780	116507	12.02				
VX0107WBL01	109213	12.02				
VX0107WBS02	167120	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.