

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1004 SAS No.: Q1004 SDG No.: Q1004

Instrument ID: MSVOA_X Calibration Date/Time: 01/07/2025 11:14

Lab File ID: VX044599.D Init. Calib. Date(s): 01/07/2025 01/07/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:06 12:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Bromoform	50.000	50.12	-0.2480011	20