

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1004

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX044607.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0107WBSD02	Chloromethane	20	20.1	ug/L	101	16		65	116	20
	Vinyl chloride	20	19.4	ug/L	97	6		65	117	20
	Bromomethane	20	20.2	ug/L	101	12		58	125	20
	Chloroethane	20	19.5	ug/L	98	11		56	128	20
	Trichlorofluoromethane	20	20.3	ug/L	102	4		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/L	97	5		80	112	20
	1,1-Dichloroethene	20	18.8	ug/L	94	5		74	110	20
	Acetone	100	120	ug/L	120	9		60	125	20
	Carbon disulfide	20	18.9	ug/L	95	8		64	112	20
	Methyl tert-butyl Ether	20	21.3	ug/L	106	12		78	114	20
	Methylene Chloride	20	19.7	ug/L	99	12		72	114	20
	trans-1,2-Dichloroethene	20	19.3	ug/L	97	9		75	108	20
	1,1-Dichloroethane	20	20.4	ug/L	102	13		78	112	20
	2-Butanone	100	120	ug/L	120	9		65	122	20
	Carbon Tetrachloride	20	19.0	ug/L	95	5		77	113	20
	cis-1,2-Dichloroethene	20	19.6	ug/L	98	11		77	110	20
	Chloroform	20	20.4	ug/L	102	14		79	113	20
	1,1,1-Trichloroethane	20	21.1	ug/L	106	11		80	108	20
	Methylcyclohexane	20	19.1	ug/L	96	2		72	115	20
	Benzene	20	19.3	ug/L	97	10		82	109	20
	1,2-Dichloroethane	20	21.3	ug/L	106	11		80	115	20
	Trichloroethene	20	18.5	ug/L	93	10		77	113	20
	1,2-Dichloropropane	20	20.5	ug/L	103	13		83	111	20
	Bromodichloromethane	20	20.6	ug/L	103	10		83	110	20
	4-Methyl-2-Pentanone	100	120	ug/L	120	18	*	74	118	20
	Toluene	20	20.3	ug/L	102	10		82	110	20
	t-1,3-Dichloropropene	20	20.8	ug/L	104	7		79	110	20
	cis-1,3-Dichloropropene	20	20.4	ug/L	102	9		82	110	20
	1,1,2-Trichloroethane	20	20.7	ug/L	104	12		83	112	20
	2-Hexanone	100	120	ug/L	120	9	*	73	117	20
	Dibromochloromethane	20	20.0	ug/L	100	8		82	110	20
	Tetrachloroethene	20	18.6	ug/L	93	8		67	123	20
	Chlorobenzene	20	19.9	ug/L	100	8		82	109	20
	Ethyl Benzene	20	20.2	ug/L	101	7		83	109	20
	m/p-Xylenes	40	40.8	ug/L	102	7		82	110	20
	o-Xylene	20	20.0	ug/L	100	8		83	109	20
	Styrene	20	20.5	ug/L	103	8		80	111	20
	Bromoform	20	20.2	ug/L	101	13		79	109	20
	Isopropylbenzene	20	20.9	ug/L	104	19		83	112	20
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107	22	*	76	118	20
	1,3-Dichlorobenzene	20	20.5	ug/L	103	10		82	108	20
	1,4-Dichlorobenzene	20	19.3	ug/L	97	9		82	107	20
	1,2-Dichlorobenzene	20	20.1	ug/L	101	9		82	109	20