

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX010625

Review By	Maresh Dadoda	Review On	1/6/2025 4:46:35 PM
Supervise By	Semsettin Yesilyurt	Supervise On	1/6/2025 4:47:05 PM
SubDirectory	VX010625	HP Acquire Method	HP Processing Method 82X121124W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132420
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132421,VP132422

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044582.D	06 Jan 2025 07:55		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX044583.D	06 Jan 2025 10:04	#44 high	JC/MD	Ok,M
3	VX0106MBL01	VX0106MBL01	VX044584.D	06 Jan 2025 10:44		JC/MD	Ok
4	VX0106WBL01	VX0106WBL01	VX044585.D	06 Jan 2025 11:07		JC/MD	Ok
5	VX0106WBS01	VX0106WBS01	VX044586.D	06 Jan 2025 11:33		JC/MD	Ok,M
6	VX0106WBSD01	VX0106WBSD01	VX044587.D	06 Jan 2025 12:00		JC/MD	Ok,M
7	Q1004-05	BP-VPB-190A-GW-803	VX044588.D	06 Jan 2025 12:23	vial A+B pH<2	JC/MD	Ok
8	Q1004-06	BP-VPB-190A-GW-818	VX044589.D	06 Jan 2025 12:46	vial A+B pH<2	JC/MD	Ok
9	Q1004-01	BP-VPB-190A-TB-2024	VX044590.D	06 Jan 2025 13:09	vial A pH<2	JC/MD	Ok
10	Q1004-03	BP-VPB-190A-EB-2024	VX044591.D	06 Jan 2025 13:32	vial A pH<2	JC/MD	Ok
11	Q1004-02	VPB190A-HYD-202412	VX044592.D	06 Jan 2025 13:55	vial A pH<2	JC/MD	Ok
12	Q1004-04	BP-VPB-190A-GW-778	VX044593.D	06 Jan 2025 14:18	hit of Trichoroethene vial A pH<2	JC/MD	Not Ok
13	VSTDCCC050	VSTDCCC050EC	VX044594.D	06 Jan 2025 15:19		JC/MD	Ok,M

M : Manual Integration