

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX010725

Review By	John Carlone	Review On	1/8/2025 9:39:14 AM
Supervise By	Mahesh Dadoda	Supervise On	1/8/2025 3:17:27 PM
SubDirectory	VX010725	HP Acquire Method	HP Processing Method 82X010725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132436
Initial Calibration Stds	VP132441,VP132442,VP132443,VP132444,VP132445,VP132446
CCC	VP132437
Internal Standard/PEM	
ICV/I.BLK	VP132447
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044595.D	07 Jan 2025 09:14		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX044596.D	07 Jan 2025 10:06		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX044597.D	07 Jan 2025 10:29	Comp.#71 is on Quadratic Regression	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX044598.D	07 Jan 2025 10:51		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX044599.D	07 Jan 2025 11:14		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX044600.D	07 Jan 2025 11:37		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX044601.D	07 Jan 2025 12:00		JC/MD	Ok,M
8	IBLK	IBLK	VX044602.D	07 Jan 2025 12:23		JC/MD	Ok,M
9	VSTDICV050	ICVVX010725	VX044603.D	07 Jan 2025 15:34		JC/MD	Ok,M
10	VX0107WBL01	VX0107WBL01	VX044604.D	07 Jan 2025 16:06		JC/MD	Ok
11	VX0107WBS01	VX0107WBS01	VX044605.D	07 Jan 2025 16:29	Recovery Fail	JC/MD	Not Ok
12	VX0107WBS02	VX0107WBS02	VX044606.D	07 Jan 2025 16:57		JC/MD	Ok,M
13	VX0107WBSD02	VX0107WBSD02	VX044607.D	07 Jan 2025 17:19		JC/MD	Ok,M
14	Q1004-04	BP-VPB-190A-GW-778	VX044608.D	07 Jan 2025 17:42	vial B pH<2	JC/MD	Ok
15	Q1022-01	BP-VPB-190A-TB-2025	VX044609.D	07 Jan 2025 18:05	vial A pH<2 TB	JC/MD	Ok
16	Q1022-03	BP-VPB-190A-GW-838	VX044610.D	07 Jan 2025 18:28	vial A+B pH<2	JC/MD	Ok
17	Q1022-04	BP-VPB-190A-GW-858	VX044611.D	07 Jan 2025 18:51	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX010725

Review By	John Carlone	Review On	1/8/2025 9:39:14 AM
Supervise By	Maresh Dadoda	Supervise On	1/8/2025 3:17:27 PM
SubDirectory	VX010725	HP Acquire Method	HP Processing Method 82X010725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132436
Initial Calibration Stds	VP132441,VP132442,VP132443,VP132444,VP132445,VP132446
CCC	VP132437
Internal Standard/PEM	
ICV/I.BLK	VP132447
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q1022-02	BP-VPB-190A-DUP-20	VX044612.D	07 Jan 2025 19:14	vial A pH<2	JC/MD	Ok
19	VSTDCCC050	VSTDCCC050EC	VX044613.D	07 Jan 2025 19:37		JC/MD	Ok,M

M : Manual Integration