



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
Data File : VX044614.D
Acq On : 08 Jan 2025 09:29
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Title : SW846 8260
Last Update : Wed Jan 08 03:57:12 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	5416	PASS
75	95	30	60	53.8	14899	PASS
95	95	100	100	100.0	27669	PASS
96	95	5	9	6.3	1739	PASS
173	174	0.00	2	1.1	225	PASS
174	95	50	100	70.9	19612	PASS
175	174	5	9	7.3	1423	PASS
176	174	95	101	98.2	19254	PASS
177	176	5	9	5.8	1115	PASS