

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1006 SAS No.: Q1006 SDG NO.: Q1006
 Lab File ID: VX044615.D Date Analyzed: 01/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	141072	12.018				
UPPER LIMIT	282144	12.518				
LOWER LIMIT	70536	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	133413	12.02				
Storage-Blank-WATER-REF-5	140600	12.02				
Storage-Blank-WATER-REF-4	129243	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	153317	12.02				
VX0108WBL01	150386	12.02				
VX0108WBS01	130106	12.02				
VX0108WBSD01	149061	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.