

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1012 SAS No.: Q1012 SDG No.: Q1012

Instrument ID: MSVOA_N Calibration Date/Time: 01/07/2025 10:30

Lab File ID: VN085390.D Init. Calib. Date(s): 12/20/2024 12/20/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:05 11:42

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.646	1.717	0.5	4.31	
Toluene	1.843	1.936	0.4	5.05	
Ethyl Benzene	1.939	1.912	0.1	-1.39	
m/p-Xylenes	0.728	0.769	0.3	5.63	
o-Xylene	0.693	0.714	0.3	3.03	
1,2-Dichloroethane-d4	2.639	2.621	0.01	-0.68	
Toluene-d8	1.465	1.505	0.01	2.73	
4-Bromofluorobenzene	0.486	0.490	0.2	0.82	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.